



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|---|-----------------------------------|---------------------------------------|
| ALCO SENSOR IV SN 099361 | PRINTER SN 097 3584 386 | DATE OF INSPECTION 3-7-2016 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 106 N 3rd St Garden City, MO 64747 | | TIME OF INSPECTION 2220 |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER **Guth** LOT # **15050** EXP. DATE **3-9-17**

SIMULATOR TEMPERATURE (34°C ± 0.2°C) **34.0°C** SIMULATOR SN **SD2314** SIMULATOR EXP DATE **11-3-16**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| | | |
|----------------------|----------------------|----------------------|
| TEST 1 = .101 | TEST 2 = .101 | TEST 3 = .102 |
|----------------------|----------------------|----------------------|

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | | | | | | | |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTOR'S OFFICER

| | |
|---|---|
| SIGNATURE <i>Richard D Williams</i> | PRINT NAME Richard D Williams |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 250305 12-22-2017 | TELEPHONE NUMBER 316-773-8201 |

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-604-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15050 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on March 11, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1218% (w/vol) ethyl alcohol. The expiration date for this lot number is March 9, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00312

Temp Date Time 210L

Air Blank:
03/07/16 22:22 .000
Calibration Check:
20 03/07/16 22:22 .101

Subject Name
Richard Williams
Subject I.D.
712

Operator Name, I.D.
250305

Location
GCPP

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00313

Temp Date Time 210L

Air Blank:
03/07/16 22:24 .000
Calibration Check:
20 03/07/16 22:24 .101

Subject Name
Richard Williams
Subject I.D.
712

Operator Name, I.D.
250305

Location
GCPP

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00314

Temp Date Time 210L

Air Blank:
03/07/16 22:26 .000
Calibration Check:
21 03/07/16 22:26 .102

Subject Name
Richard Williams
Subject I.D.
712

Operator Name, I.D.
250305

Location
GCPP

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00315

Temp Date Time 210L

VOID: RFI
12 03/07/16 22:27

Subject Name
Richard Williams
Subject I.D.
712

Operator Name, I.D.
250305

Location
GCPP



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

RICHARD D WILLIAMS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/22/2015

NUMBER 250305

EXPIRES 12/22/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-1 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WILLIAMS, RICHARD
Permit No 250305
Date Issued 12/22/2015 Date Expires 12/22/2017