

AS IV Serial no: 099364
Version no: 532B

TEST RECORD 00417

Temp Date Time ^{s/} 210L

Air Blank:
07/05/16 09:30 .000
Subject Test: Man
20 07/05/16 09:30 .105

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 099364
Version no: 532B

TEST RECORD 00418

Temp Date Time ^{s/} 210L

VOID: RFI
12 07/05/16 09:34

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 099364
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00415

Temp Date Time ^{s/} 210L

Air Blank:
07/05/16 08:58 .000
Calibration Check:
-20 07/05/16 08:58 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 099364
Version no: 532B

TEST RECORD 00416

Temp Date Time ^{s/} 210L

Air Blank:
07/05/16 09:26 .000
Calibration Check:
19 07/05/16 09:26 .105

Subject Name

Subject I.D.

Operator Name, I.D.

Location



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 15-Sep-2015

Lot # AG525701 Model 108cadd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
14-Sep-2017	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance


Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2015.09.15 16:04:10 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: _____


Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

MICHAEL RICHARDSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/31/2016

NUMBER 260168

EXPIRES 3/31/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator RICHARDSON, MICHAEL
Permit No 260168
Date Issued 3/31/2016 Date Expires 3/31/2018