



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 8:54 am, Apr 11, 2016

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 099365	PRINTER SN 097.3584.346	DATE OF INSPECTION 04/09/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 147 South Main St. Laurie, MO. 65038		TIME OF INSPECTION 10:18 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories LOT # 15120 EXP. DATE 04/29/2017
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34°C SIMULATOR SN DR6929 SIMULATOR EXP DATE 09/01/2016

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ 0.098	TEST 2 ➡ 0.100	TEST 3 ➡ 0.100
----------------	----------------	----------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Scott DeWayne Craig
TYPE II PERMIT NUMBER/EXPIRATION DATE 250048 02/20/2017	TELEPHONE NUMBER (573) 374-4871

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 00465

Temp Date Time 210L

Air Blank:
04/09/16 22:18 .000
Subject Test: Auto
19 04/09/16 22:18 .000

Subject Name

Monthly Maintenance

Subject I.D.

Self Test

Operator Name, I.D.

Scott Craig 250048

Location

147 S. Main St.

Laurie, MO. 65038

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 00467

Temp Date Time 210L

VOIH: RFI
12 04/09/16 22:20

Subject Name

Monthly Maintenance

Subject I.D.

RFI Check

Operator Name, I.D.

Scott Craig 250048

Location

147 S. Main St.

Laurie, MO. 65038

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 00468

Temp Date Time 210L

Air Blank:
04/09/16 22:21 .000
Calibration Check:
21 04/09/16 22:21 .098

Subject Name

Monthly Maintenance

Subject I.D.

Accuracy Check #1

Operator Name, I.D.

Scott Craig 250048

Location

147 S. Main St.

Laurie, MO. 65038

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 00469

Temp Date Time 210L

Air Blank:
04/09/16 22:23 .000
Calibration Check:
21 04/09/16 22:23 .100

Subject Name

Monthly Maintenance

Subject I.D.

Accuracy Check #2

Operator Name, I.D.

Scott Craig 250048

Location

147 S. Main St.

Laurie, MO. 65038

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 00470

Temp Date Time 210L

Air Blank:
04/09/16 22:24 .000
Calibration Check:
22 04/09/16 22:24 .100

Subject Name

Monthly Maintenance

Subject I.D.

Accuracy Check #3

Operator Name, I.D.

Scott Craig 250048

Location

147 S. Main St.

Laurie, MO. 65038



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
SCOTT D CRAIG

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/20/2015
NUMBER 250048
EXPIRES 2/20/2017



DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (8-10)

LAB-4 (R8-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **CRAIG, SCOTT**
Permit No **250048**
Date issued **2/20/2015** Date Expires **2/20/2017**



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **15120** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 4, 2015**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 29, 2017** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.