



AS IV Serial no: 899365  
Version no: 532B

TEST RECORD 00489

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/10/16 14:43 .000  
Subject Test: Auto  
21 06/10/16 14:43 .000

Subject Name

Blank Test

Subject I.D.

Monthly Maintenance

Operator Name, I.D.

Scott Craig 250048

Location

147 South Main St.

Laurie, MO 65038

AS IV Serial no: 899365  
Version no: 532B

TEST RECORD 00490

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 06/10/16 14:44

Subject Name

RFI Test

Subject I.D.

Monthly Maintenance

Operator Name, I.D.

Scott Craig 250048

Location

147 South Main St.

Laurie, MO 65038

AS IV Serial no: 899365  
Version no: 532B

TEST RECORD 00491

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/10/16 14:45 .000  
Calibration Check:  
22 06/10/16 14:45 .098

Subject Name

Accuracy Test #1

Subject I.D.

Monthly Maintenance

Operator Name, I.D.

Scott Craig 250048

Location

147 South Main St.

Laurie, MO 65038

AS IV Serial no: 899365  
Version no: 532B

TEST RECORD 00492

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/10/16 14:47 .000  
Calibration Check:  
23 06/10/16 14:47 .099

Subject Name

Accuracy Test #2

Subject I.D.

Monthly Maintenance

Operator Name, I.D.

Scott Craig 250048

Location

147 South Main St.

Laurie, MO 65038

AS IV Serial no: 899365  
Version no: 532B

TEST RECORD 00493

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/10/16 14:50 .000  
Calibration Check:  
24 06/10/16 14:50 .098

Subject Name

Accuracy Test #3

Subject I.D.

Monthly Maintenance

Operator Name, I.D.

Scott Craig 250048

Location

147 South Main St.

Laurie, MO 65038



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

PERMIT  
TYPE II  
SCOTT D CRAIG

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/20/2015

NUMBER 250048

EXPIRES 2/20/2017

  
\_\_\_\_\_  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

  
\_\_\_\_\_  
acting director  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 880-0771 (9-10)

LAB-4 (98-10)

 STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath from of expired air in Missouri.



Operator **CRAIG, SCOTT**  
Permit No **250048**  
Date issued **2/20/2015** Date Expires **2/20/2017**



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **16040** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **January 22, 2016**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1213%** (w/vol) ethyl alcohol. The expiration date for this lot number is **January 20, 2018** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*