

AS IV Serial no: 099366
Version no: 532B

TEST RECORD 00161

Temp	Date	Time	s/	210L
------	------	------	----	------

VOID: RFI
12 05/12/16 16:34

Subject Name

RFJ-TEST

Subject I.D.

Operator Name, I.D.

THEMANN 240401

Location

4608 OAKRIDGE BLVD.

ST. LOUIS, MO

AS IV Serial no: 099366
Version no: 532B

TEST RECORD 00158

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
05/12/16 16:29 .000
Calibration Check:
24 05/12/16 16:29 .100

Subject Name

TEST - 1

Subject I.D.

Operator Name, I.D.

THEMANN 240401

Location

4608 OAKRIDGE BLVD.

ST. LOUIS, MO

AS IV Serial no: 099366
Version no: 532B

TEST RECORD 00160

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
05/12/16 16:32 .000
Calibration Check:
25 05/12/16 16:32 .100

Subject Name

TEST - 3

Subject I.D.

Operator Name, I.D.

THEMANN 240401

Location

4608 OAKRIDGE BLVD.

ST. LOUIS, MO

AS IV Serial no: 099366
Version no: 532B

TEST RECORD 00159

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
05/12/16 16:31 .000
Calibration Check:
25 05/12/16 16:31 .101

Subject Name

TEST - 2

Subject I.D.

Operator Name, I.D.

THEMANN 240401

Location

4608 OAKRIDGE BLVD.

ST. LOUIS, MO



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 6, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1213% (w/vol) ethyl alcohol. The expiration date for this lot number is August 5, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

SHAWN J THIEMANN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/13/2014

NUMBER 240401

EXPIRES 11/13/2016

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator THIEMANN, SHAWN
 Permit No 240401
 Date Issued 11/13/2014 Date Expires 11/13/2016