RECEIVED
By Carol Day at 10:40 am, Feb 04, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
100288	08C.3527.174	02/01/2016

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
9928 GRAVOIS ROAD ST. LOUIS	12:37 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☐ SIMULATOR SOLUTION ☑ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER INTOXIMETERS LOT # AG533901 EXP. DATE 08/05/2017

☐ SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ 0.099	TEST 2 ☛ 0.099	TEST 3 ☛ 0.099

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	0	(.05-.09)	1	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE	PRINT NAME
▸ 3086	DONALD M. JACQUIN
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
250318 12/28/2017	(314) 677-9921

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

ALCO-SENSOR IV MONTHLY MAINTENANCE REPORT **FEBRUARY 2016**

UNIT #100288 **PRINTER #08C.3527.174**

AS IV Serial no: 100288
Version no: 532B

TEST RECORD 00125

Temp	Date	Time	g/210L

Air Blank:
02/01/16 12:37 .000
Subject Test: Auto
20 02/01/16 12:37 .000

Subject Name
SOBER TEST
Subject I.D.
N/A
Operator Name, I.D.
JACQUIN, 250318
Location
9928 GRAVOIS RD

PCT 3

AS IV Serial no: 100288
Version no: 532B

TEST RECORD 00126

Temp	Date	Time	g/210L

Air Blank:
02/01/16 12:38 .000
Calibration Check:
21 02/01/16 12:38 .099

Subject Name
TEST #1
Subject I.D.
N/A
Operator Name, I.D.
JACQUIN, 250318
Location
9928 GRAVOIS RD

PCT 3

AS IV Serial no: 100288
Version no: 532B

TEST RECORD 00127

Temp	Date	Time	g/210L

Air Blank:
02/01/16 12:40 .000
Calibration Check:
22 02/01/16 12:40 .099

Subject Name
TEST #2
Subject I.D.
N/A
Operator Name, I.D.
JACQUIN, 250318
Location
9928 GRAVOIS RD

PCT 3

AS IV Serial no: 100288
Version no: 532B

TEST RECORD 00128

Temp	Date	Time	g/210L

Air Blank:
02/01/16 12:41 .000
Calibration Check:
22 02/01/16 12:41 .099

Subject Name
TEST #3
Subject I.D.
N/A
Operator Name, I.D.
JACQUIN, 250318
Location
9928 GRAVOIS RD

PCT 3

AS IV Serial no: 100288
Version no: 532B

TEST RECORD 00129

Temp	Date	Time	g/210L

VOID: RFI
12 02/01/16 12:42

Subject Name
RFI
Subject I.D.
N/A
Operator Name, I.D.
JACQUIN, 250318
Location
9928 GRAVOIS RD

PCT 3

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

DONALD M JACQUIN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/28/2015

NUMBER 250318

EXPIRES 12/28/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator JACQUIN, DONALD

Permit No 250318

Date Issued 12/28/2015 **Date Expires** 12/28/2017

Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 7-Dec-2015

Lot # AG533901 **Model** 30cacd

Exp. Date	Cyl. Type	Component	Certified Concentration
5-Aug-2017	30	Ethanol	0.100 ± 2% BrAC (272 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.	Concentration	Serial No.	Concentration
EB0010581	**391.8 ppm**	**EB0010603**	**392.5 ppm**
EB0010570	**259.8 ppm**	**EB0010559**	**258.9 ppm**
EB0010285	**209.0 ppm**	**EB0010595**	**208.9 ppm**
EB0010561	**103.7 ppm**	**EB0010562**	**104.9 ppm**
EB0010681	**52.22 ppm**	**EB0010579**	**52.94 ppm**

Analytical Method: **NDIR**

Digitally signed by Quality Control
Date: 2015.12.07 14:29:01 -06:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: Rod Marsala
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01