



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Carol Day at 9:50 am, Sep 07, 2016

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 100291	PRINTER SN 093.3563.019	DATE OF INSPECTION 08/28/2016
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Willow Springs Police Department, 700 W. Main St	TIME OF INSPECTION 8:16 pm
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 14220 EXP. DATE 09/24/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.1 SIMULATOR SN SD2239 SIMULATOR EXP DATE 04/01/2017

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .097

TEST 2 .097

TEST 3 .098

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME  
Wes Ellison

TYPE II PERMIT NUMBER/EXPIRATION DATE  
250084 05/11/2017

TELEPHONE NUMBER  
(417) 469-3158

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IV Serial no: 100291  
Version no: 532B

TEST RECORD 00213

Temp Date Time <sup>sw</sup> 210L

Air Blank:  
08/28/16 20:16 .000  
Subject Test: Auto  
25 08/28/16 20:16 .000

Subject Name

NON-ALCOHOL

Subject I.D.

ELLISON

Operator Name, I.D.

250084

Location

WILLOW SPRINGS PD

AS IV Serial no: 100291  
Version no: 532B

TEST RECORD 00214

Temp Date Time <sup>sw</sup> 210L

Air Blank:  
08/28/16 20:19 .000  
Calibration Check:  
25 08/28/16 20:19 .097

Subject Name

TEST #1

Subject I.D.

ELLISON

Operator Name, I.D.

250084

Location

WILLOW SPRINGS PD

AS IV Serial no: 100291  
Version no: 532B

TEST RECORD 00215

Temp Date Time <sup>sw</sup> 210L

Air Blank:  
08/28/16 20:22 .000  
Calibration Check:  
26 08/28/16 20:22 .097

Subject Name

TEST #2

Subject I.D.

ELLISON

Operator Name, I.D.

250084

Location

WILLOW SPRINGS PD

AS IV Serial no: 100291  
Version no: 532B

TEST RECORD 00216

Temp Date Time <sup>sw</sup> 210L

Air Blank:  
08/28/16 20:25 .000  
Calibration Check:  
26 08/28/16 20:25 .098

Subject Name

TEST #3

Subject I.D.

ELLISON

Operator Name, I.D.

250084

Location

WILLOW SPRINGS PD

AS IV Serial no: 100291  
Version no: 532B

TEST RECORD 00217

Temp Date Time <sup>sw</sup> 210L

VOID: RFI  
12 08/28/16 20:25

Subject Name

RFI

Subject I.D.

ELLISON

Operator Name, I.D.

250084

Location

WILLOW SPRINGS PD



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**WES ELLISON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2015

NUMBER 250084

EXPIRES 5/11/2017

  
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

  
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0711 (5-12)

LAE 4 (REV 10)

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator ELLISON, WES  
 Permit No 250084  
 Date Issued 5/11/2015 Date Expires 5/11/2017



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14220 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 25, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 24, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*