



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 100292	PRINTER SN 093.3563.020	DATE OF INSPECTION 01/20/2016
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Pettis County Zone Office, Sedalia, Missouri	TIME OF INSPECTION 12:59 pm
--	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER RepCo Marketing, Inc. LOT # 14001 EXP. DATE 04/30/2016
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 33.9 SIMULATOR SN MP2203 SIMULATOR EXP DATE 07/21/2016

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .101	TEST 2 .100	TEST 3 .099
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Neil K. Johnson #1112
TYPE II PERMIT NUMBER/EXPIRATION DATE 250173 / 7/28/2017	TELEPHONE NUMBER (816) 200-9350

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

093.3563.020

AS IV Serial no: 100292
Version no: 532B

TEST RECORD 00045

Temp Date Time ^{s/} 210L

Air Blank:
01/20/16 12:59 .000
Calibration Check:
22 01/20/16 12:59 .000

Subject Name
BLANK
Subject I.D.

Operator Name, I.D.
NEIL K. JOHNSON
Location
PETTIS CO. ZONE OFFICE
SEDALIA, MO.

AS IV Serial no: 100292
Version no: 532B

TEST RECORD 00046

Temp Date Time ^{s/} 210L

Air Blank:
01/20/16 13:00 .000
Subject Test: Auto
22 01/20/16 13:00 .000

Subject Name
BLANK
Subject I.D.

Operator Name, I.D.
NEIL K. JOHNSON
Location
PETTIS CO. ZONE OFFICE
SEDALIA, MO.

AS IV Serial no: 100292
Version no: 532B

TEST RECORD 00047

Temp Date Time ^{s/} 210L

Air Blank:
01/20/16 13:03 .000
Calibration Check:
23 01/20/16 13:03 .101

Subject Name
MAINT.
Subject I.D.

Operator Name, I.D.
NEIL K. JOHNSON
Location
PETTIS CO. ZONE OFFICE
SEDALIA, MO.

AS IV Serial no: 100292
Version no: 532B

TEST RECORD 00048

Temp Date Time ^{s/} 210L

Air Blank:
01/20/16 13:05 .000
Calibration Check:
24 01/20/16 13:05 .100

Subject Name
MAINT.
Subject I.D.

Operator Name, I.D.
NEIL K. JOHNSON
Location
PETTIS CO. ZONE OFFICE
SEDALIA, MO.

AS IV Serial no: 100292
Version no: 532B

TEST RECORD 00049

Temp Date Time ^{s/} 210L

Air Blank:
01/20/16 13:08 .000
Calibration Check:
24 01/20/16 13:08 .099

Subject Name
MAINT.
Subject I.D.

Operator Name, I.D.
NEIL K. JOHNSON
Location
PETTIS CO. ZONE OFFICE
SEDALIA, MO.

AS IV Serial no: 100292
Version no: 532B

TEST RECORD 00050

Temp Date Time ^{s/} 210L

VOID: RFI
12 01/20/16 13:09

Subject Name
MAINT.
Subject I.D.

Operator Name, I.D.
NEIL K. JOHNSON
Location
PETTIS CO. ZONE OFFICE
SEDALIA, MO.