



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102456	PRINTER SN 096.3580.982	DATE OF INSPECTION 03/02/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 600 E WALNUT COLUMBIA		TIME OF INSPECTION 12:50 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER INTOXIMETERS LOT # AG505101 EXP. DATE 02/20/2017
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .101	TEST 2 ➡ .101	TEST 3 ➡ .101
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	2	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME SPIRIT STEVENS
TYPE II PERMIT NUMBER/EXPIRATION DATE 344517-11/16/17	TELEPHONE NUMBER (573) 442-6131

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 23-Feb-2015

Lot # AG505101

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
20-Feb-2017	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2015.02.23 15:38:13 -06:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 102456
Version no: 532B

TEST RECORD 00034

Temp Date Time ^{9/} 210L

Air Blank:
03/02/16 12:54 .000
Calibration Check:
19 03/02/16 12:54 .101

Subject Name

Cal 2

Subject I.D.

S. Stevens 2113

Operator Name, I.D.

CPD

Location

AS IV Serial no: 102456
Version no: 532B

TEST RECORD 00035

Temp Date Time ^{9/} 210L

Air Blank:
03/02/16 12:55 .000
Calibration Check:
19 03/02/16 12:55 .101

Subject Name

Cal 3

Subject I.D.

S. Stevens 2113

Operator Name, I.D.

CPD

Location

AS IV Serial no: 102456
Version no: 532B

TEST RECORD 00033

Temp Date Time ^{9/} 210L

Air Blank:
03/02/16 12:52 .000
Calibration Check:
18 03/02/16 12:52 .101

Subject Name

Cal 1

Subject I.D.

S. Stevens 2113

Operator Name, I.D.

CPD

Location

AS IV Serial no: 102456
Version no: 532B

TEST RECORD 00036

Temp Date Time ^{9/} 210L

VOID: RFI
12 03/02/16 12:56

Subject Name

RFI

Subject I.D.

S. Stevens 2113

Operator Name, I.D.

CPD

Location