



AS IV Serial no: 182459  
Version no: 532B

TEST RECORD 00097

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
02/01/16 13:23 .000  
Calibration Check:  
26 02/01/16 13:23 .099

Subject Name

MAINT TEST  
Subject I.D.

Operator Name, I.D.

S.D. Block 661  
Location

AS IV Serial no: 182459  
Version no: 532B

TEST RECORD 00099

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
02/01/16 13:25 .000  
Calibration Check:  
27 02/01/16 13:25 .099

Subject Name

MAINT TEST  
Subject I.D.

Operator Name, I.D.

S.D. Block 661  
Location

AS IV Serial no: 182459  
Version no: 532B

TEST RECORD 00098

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
02/01/16 13:24 .000  
Calibration Check:  
26 02/01/16 13:24 .099

Subject Name

MAINT TEST  
Subject I.D.

Operator Name, I.D.

S.D. Block 661  
Location

AS IV Serial no: 182459  
Version no: 532B

TEST RECORD 00100

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 02/01/16 13:27

Subject Name

RFI TEST  
Subject I.D.

Operator Name, I.D.

S.D. Block 661  
Location

**REPCO MARKETING INC.**3101-188 STONYBROOK DRIVE  
RALEIGH, N.C. 27604  
919-876-5480**CERTIFICATE OF ANALYSIS****MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.****LOT NUMBER: 15001****EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015  
The expiration date for this lot number is May 20, 2017 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**SHANNON D BLEDSOE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/30/2014

NUMBER 240195

EXPIRES 4/30/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator BLEDSON, SHANNON  
 Permit No 240195  
 Date Issued 4/30/2014 Date Expires 4/30/2016