

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 14001

EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014
The expiration date for this lot number is April 30, 2016 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

AS IV Serial no: 102459
Version no: 532B

TEST RECORD 00172

Temp Date Time ^{s/} 210L

Air Blank:
04/29/16 10:39 .000
Calibration:
25 04/29/16 10:39 .100

Subject Name CALIBRATION
MAINT TEST
Subject I.D.

Operator Name, I.D.
Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 102459
Version no: 532B

TEST RECORD 00173

Temp Date Time ^{s/} 210L

Air Blank:
04/29/16 10:50 .000
Calibration Check:
25 04/29/16 10:50 .100

Subject Name
MAINT TEST
Subject I.D.

Operator Name, I.D.
Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 102459
Version no: 532B

TEST RECORD 00174

Temp Date Time ^{s/} 210L

Air Blank:
04/29/16 10:52 .000
Calibration Check:
25 04/29/16 10:52 .100

Subject Name
MAINT TEST
Subject I.D.

Operator Name, I.D.
Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 102459
Version no: 532B

TEST RECORD 00175

Temp Date Time ^{s/} 210L

Air Blank:
04/29/16 10:53 .000
Calibration Check:
26 04/29/16 10:53 .100

Subject Name
MAINT TEST
Subject I.D.

Operator Name, I.D.
Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 102459
Version no: 532B

TEST RECORD 00176

Temp Date Time ^{s/} 210L

VOID: RFI
12 04/29/16 10:54

Subject Name
RPI TEST
Subject I.D.

Operator Name, I.D.
Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

JIMMY L CLEVELAND

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2015

NUMBER 250155

EXPIRES 7/22/2017

MO-52 (6/21 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB 4 (RS 10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator CLEVELAND, JIMMY
 Permit No 250155
 Date Issued 7/22/2015 Date Expires 7/22/2017