



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE

RECEIVED

REPORT #7

Complete this report in duplicate at the time of the regular monthly pre-employment breathalyzer calibration. Send copy to Department of Health and Senior Services; retain original in department file.

By Carol Day at 12:37 pm, Jun 15, 2016

ALCO SENSOR IV SN 102459	PRINTER SN 096.3580.930	DATE OF INSPECTION 06/14/2016
LOCATION OF INSTRUMENT (STREET AND CITY) Troop F Headquarters, Jefferson City, Missouri		TIME OF INSPECTION 11:37 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER RepCo Marketing, Inc. LOT # 15001 EXP. DATE 05/20/2017
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIMULATOR SN MP2123 SIMULATOR EXP DATE 04/28/2017
- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .103	TEST 2 .102	TEST 3 .102
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	1	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument is operating within DHSS specifications.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Richard R. Dowd
TYPE II PERMIT NUMBER/EXPIRATION DATE 260226 / 06/01/2018	TELEPHONE NUMBER (573) 751-1000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT #7

RECEIVED

By Carol Day at 12:36 pm, Jun 15, 2016

Complete this report in duplicate at the time of the regular monthly permit inspection. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102459	PRINTER SN 096.3580.930	DATE OF INSPECTION 06/14/2016
LOCATION OF INSTRUMENT (STREET AND CITY) Troop F Headquarters, Jefferson City, Missouri		TIME OF INSPECTION 11:37 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER RepCo Marketing, Inc. LOT # 15001 EXP. DATE 05/20/2017
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .103	TEST 2 ← .102	TEST 3 ← .102
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	1	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument is operating within DHSS specifications.

INSPECTING OFFICER

SIGNATURE <i>Richard R. Dowd</i>	PRINT NAME Richard R. Dowd
TYPE II PERMIT NUMBER/EXPIRATION DATE 260226 / 06/01/2018	TELEPHONE NUMBER (573) 751-1000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 102459
Version no: 532B

TEST RECORD 00210

Temp Date Time ^{9/} 210L

Air Blank:
06/14/16 22:39 .000
Calibration Check:
22 06/14/16 22:39 .102

Subject Name

ACCURACY TEST #2

Subject I.D.

TPR RR Dowd #932

Operator Name, I.D.

TROOP F H.Q.

Location

Temp Date Time 210L

Air Blank:
06/14/16 22:37 .000
Calibration Check:
21 06/14/16 22:37 .103

Subject Name

ACCURACY TEST #1

Subject I.D. #932

TPR RR Dowd

Operator Name, I.D.

TROOP F H.Q.

Location

AS IV Serial no: 102459
Version no: 532B

TEST RECORD 00212

Temp Date Time ^{9/} 210L

VOID: RFI
12 06/14/16 22:46

Subject Name

RFI TEST

Subject I.D.

TPR RR Dowd #932

Operator Name, I.D.

TROOP F H.Q.

Location

AS IV Serial no: 102459
Version no: 532B

TEST RECORD 00211

Temp Date Time ^{9/} 210L

Air Blank:
06/14/16 22:42 .000
Calibration Check:
23 06/14/16 22:42 .102

Subject Name

ACCURACY TEST #3

Subject I.D.

TPR RR Dowd #932

Operator Name, I.D.

TROOP F H.Q.

Location

REPCO MARKETING INC.3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480**CERTIFICATE OF ANALYSIS****MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.****LOT NUMBER: 15001****EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015
The expiration date for this lot number is May 20, 2017 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.