



RECEIVED

By Carol Day at 8:45 am, May 03, 2016

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <u>102467</u>	PRINTER SN <u>093.3578.247</u>	DATE OF INSPECTION <u>05-02-2016</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>201 North Street, Waynesville, MO 65583</u>		TIME OF INSPECTION <u>1400 / 2:00 pm</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Repco Marketing LOT # 15001 EXP. DATE May 20, 2017

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD 2283 SIMULATOR EXP DATE 11-05-16

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .098 TEST 2 • .098 TEST 3 • .098

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE <u>Kyle Prock</u>	PRINT NAME <u>Kyle Prock</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>250310 12/28/2017</u>	TELEPHONE NUMBER <u>573-774-2414</u>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 182467
Version no: 532B

TEST RECORD 00091

Temp Date Time 210L

Air Blank:
05/02/16 14:07 .000
Calibration Check:
27 05/02/16 14:07 .098

Subject Name
Main Test

Subject I.D.
Rock 706

Operator Name, I.D.

201 North St
Location

AS IV Serial no: 182467
Version no: 532B

TEST RECORD 00093

Temp Date Time 210L

Air Blank:
05/02/16 14:11 .000
Calibration Check:
27 05/02/16 14:11 .098

Subject Name
Main Test

Subject I.D.
Rock 706

Operator Name, I.D.

201 North St
Location

AS IV Serial no: 182467
Version no: 532B

TEST RECORD 00092

Temp Date Time 210L

Air Blank:
05/02/16 14:09 .000
Calibration Check:
27 05/02/16 14:09 .098

Subject Name
Main Test

Subject I.D.
Rock 706

Operator Name, I.D.

201 North St
Location

AS IV Serial no: 182467
Version no: 532B

TEST RECORD 00094

Temp Date Time 210L

VOID: RFI
12 05/02/16 14:12

Subject Name

Main test

Subject I.D.

Rock 706

Operator Name, I.D.

Location
201 North St

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 15001

EXPIRATION DATE: May 20, 2017 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015
The expiration date for this lot number is May 20, 2017 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
KYLE R PROCK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/28/2015

NUMBER 250310

EXPIRES 12/28/2017

MO 580-0771 (6-10)


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator PROCK, KYLE
 Permit No 250310
 Date Issued 12/28/2015 Date Expires 12/28/2017