



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102468	PRINTER SN 095-3583-748	DATE OF INSPECTION 01/01/2016
LOCATION OF INSTRUMENT (STREET AND CITY) City of Charlack City Hall 8401 MIDLAND BLVD. CHARLACK, MO. 63114		TIME OF INSPECTION 9:08 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION                       COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LABORATORIES      LOT # 14220      EXP. DATE 09/24/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0      SIMULATOR SN SD2760      SIMULATOR EXP DATE 01/15/2016

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 $\rightarrow$ 0.096	TEST 2 $\rightarrow$ 0.096	TEST 3 $\rightarrow$ 0.096
----------------------------	----------------------------	----------------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	2	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME P.O. ERIC SONTHEIMER #273
TYPE II PERMIT NUMBER/EXPIRATION DATE 250009 exp 01/02/2017	TELEPHONE NUMBER (314) 427-4715

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IV Serial no: 102468  
Version no: 532B

AS IV Serial no: 102468  
Version no: 532B

TEST RECORD 00244

TEST RECORD 00245

Temp Date Time 210L

Temp Date Time 210L

Air Blank:  
01/01/16 09:06 .000  
Calibration Check:  
22 01/01/16 09:06 .096

Air Blank:  
01/01/16 09:07 .000  
Calibration Check:  
23 01/01/16 09:07 .096

Subject Name  
SIMULATOR

Subject Name  
SIMULATOR

Operator Name, I.D.  
JC 227

Operator Name, I.D.  
JC 227

Location  
CHARLACK CRT 1A4

Location  
CHARLACK CRT 1A4

AS IV Serial no: 102468  
Version no: 532B

AS IV Serial no: 102468  
Version no: 532B

TEST RECORD 00246

TEST RECORD 00247

Temp Date Time 210L

Temp Date Time 210L

Air Blank:  
01/01/16 09:09 .000  
Calibration Check:  
24 01/01/16 09:09 .096

VOID: RFI  
12 01/01/16 09:10

Subject Name  
SIMULATOR

Subject Name  
SIMULATOR

Operator Name, I.D.  
JC 227

Operator Name, I.D.  
JC 227

Location  
CHARLACK CRT 1A4

Location  
CHARLACK CRT 1A4

**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14220 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 25, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is September 24, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Certified Reference Standard lot number FN08051301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

PERMIT  
- TYPE II

ERIC C SONTHEIMER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/2/2015

NUMBER 250009

EXPIRES 1/2/2017

MO 580-0771 (8-10)


[Signature]  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY  
[Signature]  
acting director  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (HS-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcohol content in breath form of expired air in Missouri.*



Operator **SONTHEIMER, ERIC**  
Permit No **250009**  
Date Issued **1/2/2015** Date Expires **1/2/2017**