



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 4, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is April 29, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
RYAN DEVOST

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **4/29/2015**

NUMBER **250077**


EXPIRES **4/29/2017**


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO-387 (11/13/13)

(46-4-915.15)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **DEVOST, RYAN**
 Permit No **250077**
 Date Issued **4/29/2015** Date Expires **4/29/2017**

AS IV Serial no: 102472
Version no: 532B

TEST RECORD 00717

Temp Date Time ^{s/} 210L

Air Blank:
03/03/16 21:31 .000
Calibration Check:
24 03/03/16 21:31 .100

Subject Name

TEST #1

Subject I.D.

Operator Name, I.D.

Location

GCSO HQ

Ryan Devost

AS IV Serial no: 102472
Version no: 532B

TEST RECORD 00718

Temp Date Time ^{s/} 210L

Air Blank:
03/03/16 21:32 .000
Calibration Check:
25 03/03/16 21:32 .100

Subject Name

TEST #2

Subject I.D.

Operator Name, I.D.

Location

GCSO HQ

Ryan Devost

AS IV Serial no: 102472
Version no: 532B

TEST RECORD 00719

Temp Date Time ^{s/} 210L

Air Blank:
03/03/16 21:34 .000
Calibration Check:
26 03/03/16 21:34 .099

Subject Name

TEST #3

Subject I.D.

Operator Name, I.D.

Location

GCSO HQ

Ryan Devost

PERFORMED ON:

03/03/2016 AT

GCSO HQ by

RYAN DEVOST Type II

PERMIT # 250077

AS-IV instrument #102472

AS IV Serial no: 182472
Version no: 532B

TEST RECORD 00715

Temp Date Time ^{9/} 210L

Air Blank:
03/03/16 21:28 .000
Calibration Check:
22 03/03/16 21:28 .000

Subject Name

SOBER SAMPLE
Subject I.D.

Operator Name, I.D.

Location

GC50 HQ

Ray Deant

AS IV Serial no: 182472
Version no: 532B

TEST RECORD 00716

Temp Date Time ^{9/} 210L

VOID: RFI
12 03/03/16 21:29

Subject Name

RFI TEST
Subject I.D.

Operator Name, I.D.

Location

GC50 HQ

Ray Deant