



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 10:49 am, May 05, 2016

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102525	PRINTER SN 84.9324.159	DATE OF INSPECTION 04/18/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 8374 Midland Blvd, Vinita Park, MO 63114		TIME OF INSPECTION 5:53 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 14220 EXP. DATE 09/24/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD2294 SIMULATOR EXP DATE 01/21/2017

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .097

TEST 2 • .099

TEST 3 • .098

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE <i>LT J Martin 220</i>	PRINT NAME Lt. J. Martin DSN: 220
TYPE II PERMIT NUMBER/EXPIRATION DATE 250089/ 05/11/2017	TELEPHONE NUMBER (314) 428-7373

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14220 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 25, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is September 24, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- 0.2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 102525
Version no: 532B

TEST RECORD 00342

Temp Date Time 210L
9/

Air Blank:
04/18/16 05:56 .000
Calibration Check:
26 04/18/16 05:56 .098

Subject Name

John Doe

Subject I.D.

123 DRIVE

Operator Name, I.D.

LT. J. MARTIN 220

Location

8374 MIDLAND Blvd

AS IV Serial no: 102525
Version no: 532B

TEST RECORD 00341

Temp Date Time 210L
9/

Air Blank:
04/18/16 05:55 .000
Calibration Check:
26 04/18/16 05:55 .099

Subject Name

John Doe

Subject I.D.

123 Drive

Operator Name, I.D.

LT. J. MARTIN 220

Location

8374 MIDLAND

AS IV Serial no: 102525
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00340

Temp Date Time 210L
9/

Air Blank:
04/18/16 05:53 .000
Calibration Check:
24 04/18/16 05:53 .097

Subject Name

John Doe

Subject I.D.

123 Drive

Operator Name, I.D.

LT. J. MARTIN 220

Location

8374 MIDLAND

AS IV Serial no: 102525
Version no: 532B

TEST RECORD 00343

Temp Date Time 210L
9/

VOID: RFI
12 04/18/16 05:58

Subject Name

John Doe

Subject I.D.

123 Drive

Operator Name, I.D.

LT. J. MARTIN 220

Location

8374. MIDLAND



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JESSE MARTIN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **5/11/2015**

NUMBER **250089**

EXPIRES **5/11/2017**

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

HD-592 (3-15) (2)

LAB 8-10-15

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **MARTIN, JESSE**
 Permit No **250089**
 Date Issued **5/11/2015** Date Expires **5/11/2017**