



AS IV Serial no: 107460  
Version no: 532B

TEST RECORD 00322

Temp	Date	Time	s/210L
------	------	------	--------

Air Blank:  
02/15/16 17:11 .000  
Calibration Check:  
24 02/15/16 17:11 .102

Subject Name

Subject I.D.

Operator Name, I.D.

John J Coney 118

Location

Marshall PD

461 W. Arrow St

AS IV Serial no: 107460  
Version no: 532B

TEST RECORD 00321

Temp	Date	Time	s/210L
------	------	------	--------

Air Blank:  
02/15/16 17:09 .000  
Calibration Check:  
23 02/15/16 17:09 .102

Subject Name

Subject I.D.

Operator Name, I.D.

John J. Coney 118

Location

Marshall PD

461 W. Arrow St

AS IV Serial no: 107460  
Version no: 532B

TEST RECORD 00320

Temp	Date	Time	s/210L
------	------	------	--------

Air Blank:  
02/15/16 17:08 .000  
Calibration Check:  
23 02/15/16 17:08 .102

Subject Name

Subject I.D.

Operator Name, I.D.

John J Coney 118

Location

Marshall PD

461 W Arrow St

AS IV Serial no: 107460  
Version no: 532B

TEST RECORD 00323

Temp	Date	Time	s/210L
------	------	------	--------

VOID: RTI  
12 02/15/16 17:12

Subject Name

Subject I.D.

Operator Name, I.D.

John J. Coney 118

Location

Marshall PD

461 W. Arrow St



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

PERMIT  
 TYPE II  
 JOHN J CONEY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of section 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/30/2014

NUMBER 240196

EXPIRES 4/30/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R5-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator CONEY, JOHN  
 Permit No 240196  
 Date Issued 4/30/2014 Date Expires 4/30/2016



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 4, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is April 29, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



**Missouri Department of Health and Senior Services**

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

Gail Vasterling  
Director



Jeremiah W. (Jay) Nixon  
Governor

**Missouri Department of Health and Senior Services Breath Alcohol Program**

***SIMULATOR CALIBRATION REPORT***

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

**SIMULATOR INFORMATION**

Agency: MARSHALL PD  
Serial Number: SD2277  
Manufacturer: Guth  
Model Number: 10-4D

**CALIBRATION RESULTS**

<u>Reference Temperature</u>	<u>Simulator Temperature</u>
33.98	34.0

This calibration was performed with NIST-Traceable Thermometer SN: 094948

This simulator was tested by: DRL

This testing was performed: 10/07/2015

This certification expires: 10/07/2016

Signature of certifying DHSS Scientist: 

Name of certifying DHSS Scientist: Ellen Strawsine