

RECEIVED

By Carol Day at 1:19 pm, Aug 05, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107460	PRINTER SN 099.3586.711	DATE OF INSPECTION 08/04/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 461 W. Arrow St., Marshall		TIME OF INSPECTION 3:48 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER Guth LOT # 15120 EXP. DATE 04/29/2017 SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2277 SIMULATOR EXP DATE 10/07/2016 CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .101

TEST 2 ← .100

TEST 3 ← .100

 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0 | (0-.04) 0 | (.05-.09) 0 | (.10-.14) 0 | (.15-.19) 0 | (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

N/A

INSPECTING OFFICER

SIGNATURE

▶

PRINT NAME

John J. Coney

TYPE II PERMIT NUMBER/EXPIRATION DATE

260206 / 05-02-2018

TELEPHONE NUMBER

(660) 886-7411

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 107460
Version no: 532B

TEST RECORD 00376

Temp Date Time ^{s/} 210L

Air Blank:
08/04/16 15:48 .000
Calibration Check:
24 08/04/16 15:48 .101

Subject Name

Subject I.D.

Operator Name, I.D.

John J Conroy #118
Location

Marshall Police Dept

AS IV Serial no: 107460
Version no: 532B

TEST RECORD 00377

Temp Date Time ^{s/} 210L

Air Blank:
08/04/16 15:50 .000
Calibration Check:
24 08/04/16 15:50 .100

Subject Name

Subject I.D.

Operator Name, I.D.

John J Conroy #118
Location

Marshall Police Dept

AS IV Serial no: 107460
Version no: 532B

TEST RECORD 00379

Temp Date Time ^{s/} 210L

Air Blank:
08/04/16 15:57 .000
Calibration Check:
26 08/04/16 15:57 .100

Subject Name

Subject I.D.

Operator Name, I.D.

John J Conroy #118
Location

Marshall Police Department

AS IV Serial no: 107460
Version no: 532B

TEST RECORD 00380

Temp Date Time ^{s/} 210L

VOID: RFI
12 08/04/16 15:58

Subject Name

Subject I.D.

Operator Name, I.D.

John J Conroy #118
Location

Marshall Police Dept



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

Gail Vasterling
Director



Jeremiah W. (Jay) Nixon
Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

SIMULATOR CALIBRATION REPORT

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

SIMULATOR INFORMATION

Agency:	<u>MARSHALL PD</u>
Serial Number:	<u>SD2277</u>
Manufacturer:	<u>Guth</u>
Model Number:	<u>10-4D</u>

CALIBRATION RESULTS

<u>Reference Temperature</u>	<u>Simulator Temperature</u>
33.98	34.0

This calibration was performed with NIST-Traceable Thermometer SN: 094948

This simulator was tested by: DRL

This testing was performed: 10/07/2015

This certification expires: 10/07/2016

Signature of certifying DHSS Scientist: 

Name of certifying DHSS Scientist: Ellen Strawsine



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 4, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is April 29, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JOHN J CONEY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/2/2016

NUMBER 260206

EXPIRES 5/2/2018

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **CONEY, JOHN**
 Permit No **260206**
 Date Issued **5/2/2016** Date Expires **5/2/2018**