



**REPCO MARKETING INC.**3101-188 STONYBROOK DRIVE  
RALEIGH, N.C. 27604  
919-876-5480**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**  
**LOT NUMBER: 15001**  
**EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015  
The expiration date for this lot number is May 20, 2017 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

AS IV Serial no: 107971  
Version no: 532B

TEST RECORD 00131

Temp Date Time 210L<sup>a/</sup>

Air Blank:  
07/02/16 19:48 .000  
Calibration Check:  
21 07/02/16 19:48 .100

Subject Name

MAINT TEST

Subject I.D.

Operator Name, I.D.

S.D. Blocker 661

Location

AS IV Serial no: 107971  
Version no: 532B

TEST RECORD 00133

Temp Date Time 210L<sup>a/</sup>

Air Blank:  
07/02/16 19:51 .000  
Calibration Check:  
23 07/02/16 19:51 .099

Subject Name

MAINT TEST

Subject I.D.

Operator Name, I.D.

S.D. Blocker 661

Location

AS IV Serial no: 107971  
Version no: 532B

TEST RECORD 00132

Temp Date Time 210L<sup>a/</sup>

Air Blank:  
07/02/16 19:50 .000  
Calibration Check:  
22 07/02/16 19:50 .099

Subject Name

MAINT TEST

Subject I.D.

Operator Name, I.D.

S.D. Blocker 661

Location

AS IV Serial no: 107971  
Version no: 532B

TEST RECORD 00134

Temp Date Time 210L<sup>a/</sup>

VOID: RFI  
12 07/02/16 19:52

Subject Name

RFI CHECK

Subject I.D.

Operator Name, I.D.

S.D. Blocker 661

Location



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**SHANNON D BLEDSOE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2016

NUMBER 260130

EXPIRES 3/7/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator BLEDSON, SHANNON  
 Permit No 260130  
 Date Issued 3/7/2016 Date Expires 3/7/2018