

Boone Co

AS IV Serial no: 187972
Version no: 532B

TEST RECORD 00190

Temp Date Time ^{9/} 210L

Air Blank:
02/15/16 10:50 .000
Calibration Check:
30 02/15/16 10:50 .079

Subject Name

DAN LUCAS
Subject I.D.

250138 6/15/2017
Operator Name, I.D.

DAN LUCAS 250138
Location 6/15/2017
1200 South Holden Street

Warrensburg MO 64093

AS IV Serial no: 187972
Version no: 532B

TEST RECORD 00191

Temp Date Time ^{9/} 210L

Air Blank:
02/15/16 10:52 .000
Calibration Check:
30 02/15/16 10:52 .078

Subject Name

DAN LUCAS
Subject I.D.

250138 6/15/2017
Operator Name, I.D.

DAN LUCAS 250138
Location 6/15/2017
1200 South Holden Street

Warrensburg MO 64093

AS IV Serial no: 187972
Version no: 532B

TEST RECORD 00192

Temp Date Time ^{9/} 210L

Air Blank:
02/15/16 10:53 .000
Calibration Check:
30 02/15/16 10:53 .078

Subject Name

DAN LUCAS
Subject I.D.

250138 6/15/2017
Operator Name, I.D.

DAN LUCAS 250138
Location 6/15/2017
1200 South Holden Street

Warrensburg MO 64093

AS IV Serial no: 187972
Version no: 532B

TEST RECORD 00193

Temp Date Time ^{9/} 210L

VOID: RFI
12 02/15/16 10:54

Subject Name

DAN LUCAS
Subject I.D.

250138 6/15/2017
Operator Name, I.D.

DAN LUCAS 250138
Location 6/15/2017
1200 South Holden Street

Warrensburg MO 64093



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

DAN R LUCAS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT, INTOX 8000, INTOX EC/IR II, ASIV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/15/2015

NUMBER 250138

EXPIRES 6/15/2017

[Signature]
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 6-Aug-2014

Lot # AG421804


<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
6-Aug-2016	108	Ethanol Nitrogen	0.080 ± 0.002 BrAC (218 ppm) Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2014.08.06 17:38:54 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Analyst: 
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01