



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 3:59 pm, Mar 31, 2016

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107975	VANDALIA PD	PRINTER SN 099 3586 829	DATE OF INSPECTION 03/17/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 1200 S. HOLDEN ST. WARRENSBURG		TIME OF INSPECTION 1000	

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)	OK
<input checked="" type="checkbox"/> TEMPERATURE OF ALCO SENSOR (10°C - 40°C)	OK
<input checked="" type="checkbox"/> PRINTER WORKING PROPERLY	OK
<input checked="" type="checkbox"/> TIME AND DATE DISPLAYING PROPERLY	OK

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER INTOXIMETERS	LOT # AG602301 EXP. DATE 01/23/2018
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____	SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .098	TEST 2 .099	TEST 3 .100
--------------------	--------------------	--------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS _____	(0-.04) _____	(.05-.09) _____	(.10-.14) _____	(.15-.19) _____	(OVER .19) _____
----------------	---------------	-----------------	-----------------	-----------------	------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**REPLACED SAMPLE CONTROL CAPACITOR.
 RECALIBRATED UNIT.**

INSPECTING OFFICER	
SIGNATURE <i>Robert Welsh</i>	PRINT NAME ROBERT WELSH
TYPE II PERMIT NUMBER/EXPIRATION DATE 250122 06/03/2017	TELEPHONE NUMBER 6605434597

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 107975
Version no: 532B

TEST RECORD 00240

Temp Date Time ^{a/} 210L

Air Blank:
03/17/16 10:00 .000
Calibration:
25 03/17/16 10:00 .098

Subject Name

CALIBRATION

Subject I.D.

WELSH 250122

Operator Name, I.D.

MSC

Location

AS IV Serial no: 107975
Version no: 532B

TEST RECORD 00241

Temp Date Time ^{a/} 210L

Air Blank:
03/17/16 10:02 .000
Calibration Check:
25 03/17/16 10:02 .098

Subject Name

CAL CK 1

Subject I.D.

WELSH 250122

Operator Name, I.D.

MSC

Location

AS IV Serial no: 107975
Version no: 532B

TEST RECORD 00242

Temp Date Time ^{a/} 210L

Air Blank:
03/17/16 10:03 .000
Subject Test: Auto
26 03/17/16 10:03 .000

Subject Name

BLANK SUB. CK.

Subject I.D.

WELSH 250122

Operator Name, I.D.

MSC

Location

AS IV Serial no: 107975
Version no: 532B

TEST RECORD 00243

Temp Date Time ^{a/} 210L

Air Blank:
03/17/16 10:06 .000
Calibration Check:
26 03/17/16 10:06 .099

Subject Name

CAL CK 2

Subject I.D.

WELSH 250122

Operator Name, I.D.

MSC

Location

AS IV Serial no: 107975
Version no: 532B

TEST RECORD 00245

Temp Date Time ^{a/} 210L

VOID: RFI
12 03/17/16 10:10

Subject Name

RFI

Subject I.D.

WELSH 250122

Operator Name, I.D.

MSC

Location

AS IV Serial no: 107975
Version no: 532B

TEST RECORD 00244

Temp Date Time ^{a/} 210L

Air Blank:
03/17/16 10:08 .000
Calibration Check:
26 03/17/16 10:08 .100

Subject Name

CAL CK 3

Subject I.D.

WELSH 250122

Operator Name, I.D.

MSC

Location



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

ROBERT W WELSH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX DMT, INTOXILYZER 5000, INTOXILYZER
 8000, ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/3/2015

NUMBER 250122

EXPIRES 6/3/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WELSH, ROBERT
 Permit No 250122
 Date Issued 6/3/2015 Date Expires 6/3/2017



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 24-Jan-2016

Lot # AG602301 Model 108cacd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
23-Jan-2018	108	Ethanol	0.100 ± 2% BrAC (272 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2016.01.25 09:20:21 -06:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01