

AS IV Serial not 107983
Version not 532B

TEST RECORD 00536

Temp Date Time 210L

Air Blank:
03/04/16 01:05 .000
Calibration Check:
23 03/04/16 01:05 .098

Subject Name

Test #3

Subject I.D.

W. Iderdyke

Operator Name, I.D.

8170

Location

12 S. Water St

AS IV Serial not 107983
Version not 532B

TEST RECORD 00535

Temp Date Time 210L

Air Blank:
03/04/16 01:03 .000
Calibration Check:
22 03/04/16 01:03 .099

Subject Name

Test #2

Subject I.D.

W. Iderdyke

Operator Name, I.D.

8170

Location

12 S. Water St

AS IV Serial not 107983
Version not 532B

TEST RECORD 00534

Temp Date Time 210L

Air Blank:
03/04/16 01:01 .000
Calibration Check:
21 03/04/16 01:01 .100

Subject Name

Test #1

Subject I.D.

W. Iderdyke

Operator Name, I.D.

8170

Location

12 S. Water St

AS IV Serial not 107983
Version not 532B

TEST RECORD 00537

Temp Date Time 210L

UNIT: RPI

12 03/04/16 01:00

Subject Name

RFR Test

Subject I.D.

W. Iderdyke

Operator Name, I.D.

8170

Location

12 S. Water St



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 20-Mar-2014

Lot # AG407801

Exp. Date

19-Mar-2016

Cyl. Type

108

Component

Ethanol
Nitrogen

Certified Concentration

0.100 ± 2% BrAC (272 ppm)
Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

Concentration

391.8 ppm

259.8 ppm

209.0 ppm

103.7 ppm

52.22 ppm

Serial No.

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

Concentration

392.5 ppm

258.9 ppm

208.9 ppm

104.9 ppm

52.94 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2014.03.20 17:26:10 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

CHAD A WILDERDYKE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/30/2014

NUMBER 240221

EXPIRES 4/30/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 590-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **WILDERDYKE, CHAD**
 Permit No **240221**
 Date Issued **4/30/2014** Date Expires **4/30/2016**