

REPCO MARKETING INC.3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480**CERTIFICATE OF ANALYSIS****MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.****LOT NUMBER: 15001****EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015
The expiration date for this lot number is May 20, 2017 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.

AS IV Serial no: 107995
Version no: 532B

TEST RECORD 00128

Temp	Date	Time	a/ 210L
------	------	------	------------

Air Blank:
05/30/16 10:53 .000
Calibration Check:
22 05/30/16 10:53 .101

Subject Name

MAINT TEST

Subject I.D.

Operator Name, I.D.

S.D. Bledsoe 661

Location

AS IV Serial no: 107995
Version no: 532B

TEST RECORD 00130

Temp	Date	Time	a/ 210L
------	------	------	------------

Air Blank:
05/30/16 10:56 .000
Calibration Check:
23 05/30/16 10:56 .104

Subject Name

MAINT TEST

Subject I.D.

Operator Name, I.D.

S.D. Bledsoe 661

Location

AS IV Serial no: 107995
Version no: 532B

TEST RECORD 00129

Temp	Date	Time	a/ 210L
------	------	------	------------

Air Blank:
05/30/16 10:54 .000
Calibration Check:
22 05/30/16 10:54 .105

Subject Name

MAINT TEST

Subject I.D.

Operator Name, I.D.

S.D. Bledsoe 661

Location

AS IV Serial no: 107995
Version no: 532B

TEST RECORD 00131

Temp	Date	Time	a/ 210L
------	------	------	------------

VOID: RFI
12 05/30/16 10:57

Subject Name

RFI CHECK

Subject I.D.

Operator Name, I.D.

S.D. Bledsoe 661

Location



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

SHANNON D BLEDSOE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2016

NUMBER 260130

EXPIRES 3/7/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator BLEDSON, SHANNON
 Permit No 260130
 Date Issued 3/7/2016 Date Expires 3/7/2018