



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108262	PRINTER SN 099.3586.831	DATE OF INSPECTION 03/02/2016
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 2920 N. Shamrock, Jefferson City	TIME OF INSPECTION 2:30 pm
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER RepCo Marketing LOT # 14001 EXP DATE 04/30/2016
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 33.99 SIMULATOR SN MP2316 SIMULATOR EXP DATE 07/20/2016

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100	TEST 2 ← .100	TEST 3 ← .100
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Initial instrument inspection and calibration.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Jimmy L. Cleveland
---------------	----------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 250155/07-22-2017	TELEPHONE NUMBER (573) 526-6193
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 108262
Version no: 532B

TEST RECORD 00364

Temp Date Time ^{s/} 210L

Air Blank:
03/02/16 14:31 .000
Calibration:
23 03/02/16 14:31 .100

Subject Name

Calibration

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Shamrock

Jefferson City

AS IV Serial no: 108262
Version no: 532B

TEST RECORD 00365

Temp Date Time ^{s/} 210L

Air Blank:
03/02/16 14:37 .000
Calibration Check:
25 03/02/16 14:37 .100

Subject Name

Jim MAINT 1

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Shamrock

Jefferson City

AS IV Serial no: 108262
Version no: 532B

TEST RECORD 00366

Temp Date Time ^{s/} 210L

Air Blank:
03/02/16 14:39 .000
Calibration Check:
25 03/02/16 14:39 .100

Subject Name

MAINT 2

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Shamrock

Jefferson City

AS IV Serial no: 108262
Version no: 532B

TEST RECORD 00367

Temp Date Time ^{s/} 210L

Air Blank:
03/02/16 14:40 .000
Calibration Check:
26 03/02/16 14:40 .100

Subject Name

MAINT 3

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Shamrock

Jefferson City

AS IV Serial no: 108262
Version no: 532B

TEST RECORD 00369

Temp Date Time ^{s/} 210L

VOID: RFI
12 03/02/16 14:42

Subject Name

RFI TEST

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Shamrock

Jefferson City

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 14001

EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014
The expiration date for this lot number is April 30, 2016 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

JIMMY L CLEVELAND

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2015

NUMBER 250155

EXPIRES 7/22/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0/71 (6-10)

LAB-4 (R5-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator CLEVELAND, JIMMY
 Permit No 250155
 Date Issued 7/22/2015 Date Expires 7/22/2017