





**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 4, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is April 29, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

AS IV Serial no: 100270  
Version no: 532B

TEST RECORD 00132

Temp Date Time 210L  
s/

Air Blank:  
05/26/16 09:11 .000  
Calibration Check:  
24 05/26/16 09:11 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Wehr 306

Location

Holden PD

AS IV Serial no: 100270  
Version no: 532B

TEST RECORD 00136

Temp Date Time 210L  
s/

Air Blank:  
05/26/16 09:10 .000  
Calibration Check:  
24 05/26/16 09:10 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Wehr 306

Location

Holden PD

AS IV Serial no: 100270  
Version no: 532B

TEST RECORD 00139

Temp Date Time 210L  
s/

VOID: RFI  
12 05/26/16 09:14

Subject Name

Subject I.D.

Operator Name, I.D.

Wehr 306

Location

Holden PD

AS IV Serial no: 100270  
Version no: 532B

TEST RECORD 00138

Temp Date Time 210L  
s/

Air Blank:  
05/26/16 09:13 .000  
Calibration Check:  
25 05/26/16 09:13 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Wehr 306

Location

Holden PD



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**TOM WEHR**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field services and repairs, and operate the following breath analyzer(s).

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of breath from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **5/11/2015**

NUMBER **250101**

EXPIRES **5/11/2017**

MISSOURI 05/11/2015

DIRECTOR OF STATE ALCOHOLIC BEVERAGE CONTROL

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator **WEHR, TOM**  
 Permit No **250101**  
 Date Issued **5/11/2015** Date Expires **5/11/2017**