RECEIVED
By Carol Day at 7:53 am, Aug 08, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
108388	099.3586.815	08/06/2016

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
9928 GRAVOIS ROAD ST. LOUIS	2:35 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☐ SIMULATOR SOLUTION ☑ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER INTOXIMETERS LOT # AG533901 EXP. DATE 08/05/2017

☐ SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ 0.098	TEST 2 ☛ 0.097	TEST 3 ☛ 0.096

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
0	0	1	2	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

11 VOIDS since last MR, sent back to Safety Center for repair.

INSPECTING OFFICER

SIGNATURE	PRINT NAME
▸ Sgt D. J. 3086	DONALD M. JACQUIN
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
250318 12/28/2017	(314) 638-5550

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

ALCO-SENSOR IV MONTHLY MAINTENANCE REPORT **AUGUST 6, 2016**

UNIT #108388 **PRINTER #099.3586.815**

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00268

Temp	Date	Time	g/210L
Air Blank:			
	08/06/16	02:35	.000
Calibration Check:			
22	08/06/16	02:35	.098

Subject Name
TEST #1

Subject I.D.
N/A

Operator Name, I.D.
SOT JACQUIN, 250318

Location
9928 GRAVOIS RD

63123

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00269

Temp	Date	Time	g/210L
Air Blank:			
	08/06/16	02:37	.000
Calibration Check:			
23	08/06/16	02:37	.097

Subject Name
TEST #2

Subject I.D.
N/A

Operator Name, I.D.
SOT. JACQUIN, 250318

Location
9928 GRAVOIS RD

63123

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00270

Temp	Date	Time	g/210L
Air Blank:			
	08/06/16	02:38	.000
Calibration Check:			
25	08/06/16	02:38	.096

Subject Name
TEST #3

Subject I.D.
N/A

Operator Name, I.D.
SOT JACQUIN, 250318

Location
9928 GRAVOIS RD

63123

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00271

Temp	Date	Time	g/210L
VOID: RFI			
12	08/06/16	02:40	

Subject Name
RFI

Subject I.D.
N/A

Operator Name, I.D.
SOT JACQUIN, 250318

Location
9928 GRAVOIS RD

63123

Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 7-Dec-2015

Lot # AG533901 **Model** 30cacd

Exp. Date	Cyl. Type	Component	Certified Concentration
5-Aug-2017	30	Ethanol	0.100 ± 2% BrAC (272 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.	Concentration	Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: **NDIR**

Digitally signed by Quality Control
Date: 2015.12.07 14:29:01 -06:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

DONALD M JACQUIN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/28/2015

NUMBER 250318

EXPIRES 12/28/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator JACQUIN, DONALD
Permit No 250318
Date Issued 12/28/2015 **Date Expires** 12/28/2017