

REPCO MARKETING INC.3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480**CERTIFICATE OF ANALYSIS**

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 15001
EXPIRATION DATE: May 20, 2017 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015
The expiration date for this lot number is May 20, 2017 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

AS IV Serial no: 108391
Version no: 532B

TEST RECORD 00287

Temp Date Time 210L ^{s/}

Air Blank:
07/26/16 09:44 .000
Calibration Check:
23 07/26/16 09:44 .100

Subject Name

Accuracy Check

Subject I.D.

Test #1

Operator Name, I.D.

T.M.P. Michels

Location

Zone 4 zone office

Troop C

AS IV Serial no: 108391
Version no: 532B

TEST RECORD 00289

Temp Date Time 210L ^{s/}

VOID: RFI
12 07/26/16 09:53

Subject Name

Maint test RFI

Subject I.D.

T.M.P. Michels

Operator Name, I.D.

Location

CO4 office

AS IV Serial no: 108391
Version no: 532B

TEST RECORD 00290

Temp Date Time 210L ^{s/}

Air Blank:
07/26/16 09:58 .000
Calibration Check:
25 07/26/16 09:58 .100

Subject Name

Maint test #1

Subject I.D.

T.M.P. Michels

Operator Name, I.D.

Location

CO4 office

AS IV Serial no: 108391
Version no: 532B

TEST RECORD 00291

Temp Date Time 210L ^{s/}

Air Blank:
07/26/16 10:11 .000
Calibration Check:
24 07/26/16 10:11 .057

Subject Name

Maint Test #2

Subject I.D.

Operator Name, I.D.

T.M.P. Michels

Location

CO4 office

AS IV Serial no: 108391
Version no: 532B

TEST RECORD 00292

Temp Date Time 210L ^{s/}

Air Blank:
07/26/16 10:14 .000
Calibration Check:
25 07/26/16 10:14 .059

Subject Name

Maint Test #3

Subject I.D.

Operator Name, I.D.

T.M.P. Michels

Location

CO4 office



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

ADAM J MICHELS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/9/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 260231

EXPIRES 6/9/2018

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator MICHELS, ADAM
 Permit No 260231
 Date Issued 6/9/2016 Date Expires 6/9/2018