



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 ALCO-SENSOR IV WITH PRINTER MAINTENANCE

RECEIVED

By Carol Day at 12:38 pm, Jul 05, 2016

REPORT#7

Complete this report in duplicate at the time of the regular monthly preventive maintenance inspection or repair.
 Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 110743	PRINTER SN 95.1111.053	DATE OF INSPECTION 07-04-2016
-----------------------------	---------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 501 FARAON ST SAINT JOSEPH, MISSOURI 64501	TIME OF INSPECTION 0050
--	----------------------------

CHECKLIST: Place a mark in the box by each item it found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (1 O-C - 40-C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER	INTOXIMETERS
LOT #	AG428002
EXP. DATE	10-07-2016
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIMULATOR SN
	SIMULATOR EXP DATE

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/>	0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
<input type="checkbox"/>	0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
<input type="checkbox"/>	0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 w- .101	TEST 2 w- .102	TEST 3 w- .099
----------------	----------------	----------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	1	(OVER.19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	-----------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME BRAD KERNS
TYPE 11 PERMIT NUMBER/EXPIRATION DATE 250129 EXP. 06-08-2017	TELEPHONE NUMBER (816) 271-5359

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Intoximeters, Inc.
 2081 Cralg Road
 St. Louis, Mo 63146

Test Date: 8-Oct-2014

Lot # AG428002

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
7-Oct-2016	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010569	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2014.10.08 12:16:00 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Analyst: _____

Rod Marsala
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

APR 17 1974
APR 17 1974

970000 00000
BY

DATE TIME TIME
04/16 08:50 000
04/16 08:50 101

04/16 08:50 000
04/16 08:50 101

04/16 08:50 101
04/16 08:50 101

04/16 08:50 101
04/16 08:50 101

04/16 08:50 101
04/16 08:50 101

APR 17 1974
APR 17 1974

970000 00000
BY

DATE TIME TIME
04/16 08:50 000
04/16 08:50 101

04/16 08:50 000
04/16 08:50 101

04/16 08:50 101
04/16 08:50 101

04/16 08:50 101
04/16 08:50 101

04/16 08:50 101
04/16 08:50 101

APR 17 1974
APR 17 1974

970000 00000
BY

DATE TIME TIME
04/16 08:50 000
04/16 08:50 101

04/16 08:50 000
04/16 08:50 101

04/16 08:50 101
04/16 08:50 101

04/16 08:50 101
04/16 08:50 101

04/16 08:50 101
04/16 08:50 101

APR 17 1974
APR 17 1974

970000 00000
BY

DATE TIME TIME
04/16 08:50 000
04/16 08:50 101

04/16 08:50 000
04/16 08:50 101

04/16 08:50 101
04/16 08:50 101

04/16 08:50 101
04/16 08:50 101

04/16 08:50 101
04/16 08:50 101



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
 TYPE II

BRAD KERNS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX EC/IR II, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/8/2015

NUMBER 250129

EXPIRES 6/8/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator KERNS, BRAD
 Permit No 250129
 Date Issued 6/8/2015 Date Expires 6/8/2017