



AS IV Serial no: 111323  
Version no: 532B

TEST RECORD 00012

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
02/29/16 07:54 .000  
Calibration Check:  
26 02/29/16 07:54 .100

Subject Name  
Calibration  
Subject I.D.

Operator Name, I.D.  
Jimmy L. Cleveland  
Location  
2920 N. Shamrock  
Jefferson City

AS IV Serial no: 111323  
Version no: 532B

TEST RECORD 00013

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
02/29/16 08:03 .000  
Calibration Check:  
27 02/29/16 08:03 .099

Subject Name  
MAINT TEST  
Subject I.D.

Operator Name, I.D.  
Jimmy L. Cleveland  
Location  
2920 N. Shamrock  
Jefferson City

AS IV Serial no: 111323  
Version no: 532B

TEST RECORD 00015

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
02/29/16 08:07 .000  
Calibration Check:  
27 02/29/16 08:07 .100

Subject Name  
MAINT TEST  
Subject I.D.

Operator Name, I.D.  
Jimmy L. Cleveland  
Location  
2920 N. Shamrock  
Jefferson City

AS IV Serial no: 111323  
Version no: 532B

TEST RECORD 00016

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
02/29/16 08:12 .000  
Calibration Check:  
27 02/29/16 08:12 .099

Subject Name  
MAINT TEST  
Subject I.D.

Operator Name, I.D.  
Jimmy L. Cleveland  
Location  
2920 N. Shamrock  
Jefferson City

AS IV Serial no: 111323  
Version no: 532B

TEST RECORD 00017

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 02/29/16 08:13

Subject Name  
RFI TEST  
Subject I.D.

Operator Name, I.D.  
Jimmy L. Cleveland  
Location  
2920 N. Shamrock  
Jefferson City

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 14001**

**EXPIRATION DATE: April 30, 2016 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014  
The expiration date for this lot number is April 30, 2016 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**JIMMY L CLEVELAND**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2015

NUMBER 250155

EXPIRES 7/22/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator CLEVELAND, JIMMY  
 Permit No 250155  
 Date Issued 7/22/2015 Date Expires 7/22/2017