



RECEIVED
By Carol Day at 1:37 pm, Jul 20, 2016

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111333	PRINTER SN 099.3586.616	DATE OF INSPECTION 07/20/2016
LOCATION OF INSTRUMENT (STREET AND CITY) Troop F Headquarters, 2920 N. Shamrock Rd, Jefferson City		TIME OF INSPECTION 11:06 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER RepCO LOT # 15001 EXP. DATE 05/20/2017
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIMULATOR SN MP2123 SIMULATOR EXP DATE 04/28/2017

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <input checked="" type="checkbox"/> .101	TEST 2 <input checked="" type="checkbox"/> .100	TEST 3 <input checked="" type="checkbox"/> .102
---	---	---

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	2	(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME M.A. Dickens
TYPE II PERMIT NUMBER/EXPIRATION DATE 260207 05/03/2018	TELEPHONE NUMBER (573) 751-1000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 111333
Version no: 532B

TEST RECORD 00085

Temp Date Time ^{a/} 210L

Air Blank:
07/20/16 11:07 .000
Calibration Check:
30 07/20/16 11:07 .101

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

MA Dickens 675

Location

Troop F

AS IV Serial no: 111333
Version no: 532B

TEST RECORD 00086

Temp Date Time ^{a/} 210L

Air Blank:
07/20/16 11:11 .000
Calibration Check:
29 07/20/16 11:11 .100

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

MA Dickens 675

Location

Troop F

AS IV Serial no: 111333
Version no: 532B

TEST RECORD 00087

Temp Date Time ^{a/} 210L

Air Blank:
07/20/16 11:13 .000
Calibration Check:
29 07/20/16 11:13 .102

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

MA Dickens 675

Location

Troop F

AS IV Serial no: 111333
Version no: 532B

TEST RECORD 00088

Temp Date Time ^{a/} 210L

VOID: RFI
12 07/20/16 11:14

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

MA Dickens 675

Location

Troop F

REPCO MARKETING INC.3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480**CERTIFICATE OF ANALYSIS****MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.****LOT NUMBER: 15001****EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015
The expiration date for this lot number is May 20, 2017 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.