

SOBEE CHECK

#1

#2

AS IV Serial no: 111749
Version no: 532B

AS IV Serial no: 111749
Version no: 532B

AS IV Serial no: 111749
Version no: 532B

TEST RECORD 00089

TEST RECORD 00092

TEST RECORD 00093

Temp Date Time ^{s/} 210L

Temp Date Time ^{s/} 210L

Temp Date Time ^{s/} 210L

Air Blank:
07/19/16 19:24 .000
Subject Test: Auto
28 07/19/16 19:24 .000

Air Blank:
07/19/16 19:30 .000
Subject Test: Auto
28 07/19/16 19:30 .102

Air Blank:
07/19/16 19:31 .000
Subject Test: Auto
28 07/19/16 19:31 .102

Subject Name

Subject Name

Subject Name

LUSK-BRINEGAR

LUSK-BRINEGAR

LUSK-BRINEGAR

Subject I.D.
#588

Subject I.D.
#588

Subject I.D.
#588

Operator Name, I.D.

Operator Name, I.D.

Operator Name, I.D.

Location

Location

Location

TROOP A ZONE B

TROOP A ZONE B

TROOP A ZONE B

OFFICE

#3

AS IV Serial no: 111749
Version no: 532B

RFI

AS IV Serial no: 111749
Version no: 532B

TEST RECORD 00094

TEST RECORD 00095

Temp Date Time ^{s/} 210L

Temp Date Time ^{s/} 210L

Air Blank:
07/19/16 19:32 .000
Subject Test: Auto
28 07/19/16 19:32 .102

VOID: RFI
12 07/19/16 19:35

Subject Name

Subject Name

LUSK-BRINEGAR

RFI TEST

Subject I.D.
#588

Subject I.D.
#588

Operator Name, I.D.

Operator Name, I.D.

Location

Location

TROOP A ZONE B

TROOP A ZONE B

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 15001

EXPIRATION DATE: May 20, 2017 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015
The expiration date for this lot number is May 20, 2017 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

ELIZABETH A LUSK-BRINEGAR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/17/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 260237

EXPIRES 6/17/2018

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 560.0771 (6-10)

LAB 4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator LUSK-BRINEGAR, ELIZABETH
Permit No 260237
Date Issued 6/17/2016 **Date Expires** 6/17/2018