RECEIVED
By Carol Day at 10:41 am, Aug 30, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
111752 (Strafford PD)	09B.3589.508	08/24/2016

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
1200 South Holden Street Warrensburg	10:08 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- [x] TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- [x] PRINTER WORKING PROPERLY
- [x] TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- [ ] SIMULATOR SOLUTION
- [x] COMPRESSED ETHANOL-GAS MIXTURE

[x] STANDARD SUPPLIER Intoximeters LOT # AG535002 EXP. DATE 12/16/2017

[ ] SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

[x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [ ] 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- [x] 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ 0.077	TEST 2 ☛ 0.077	TEST 3 ☛ 0.077

[x] RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument assigned to Strafford Police Department.

INSPECTING OFFICER

SIGNATURE	PRINT NAME
▸	Dan R Lucas Jr
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
250138 06/15/2017	(660) 543-4573

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

AS IV Serial no: 111752
Version no: 532B

TEST RECORD 00004

Temp	Date	Time	g/ 210L
Air Blank:			
	08/24/16	10:08	.000
Calibration Check:			
23	08/24/16	10:08	.077

Subject Name
Test

Subject I.D.
#1

Operator Name, I.D.
Dan Lucas 250138

Location EXP: 06/15/2017
1200 S. Holden Street
Warrensburg MO 64093

AS IV Serial no: 111752
Version no: 532B

TEST RECORD 00005

Temp	Date	Time	g/ 210L
Air Blank:			
	08/24/16	10:11	.000
Calibration Check:			
24	08/24/16	10:11	.077

Subject Name
Test

Subject I.D.
#2

Operator Name, I.D.
Dan Lucas 250138

Location EXP 06/15/2017
1200 S. Holden Street
Warrensburg MO 64093

AS IV Serial no: 111752
Version no: 532B

TEST RECORD 00006

Temp	Date	Time	g/ 210L
Air Blank:			
	08/24/16	10:13	.000
Calibration Check:			
25	08/24/16	10:13	.077

Subject Name
Test

Subject I.D.
#3

Operator Name, I.D.
Dan Lucas 250138

Location EXP: 06/15/2017
1200 S. Holden Street
Warrensburg MO 64093

AS IV Serial no: 111752
Version no: 532B

TEST RECORD 00007

Temp	Date	Time	g/ 210L
VOID: RFI			
12	08/24/16	10:14	

Subject Name
RFI

Subject I.D.
Test

Operator Name, I.D.
Dan Lucas 250138

Location EXP 06/15/2017
1200 S. Holden Street
Warrensburg MO 64093

Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 21-Dec-2015

Lot # AG535002 **Model** 108cacd

Exp. Date	Cyl. Type	Component	Certified Concentration
16-Dec-2017	108	Ethanol	0.080 ± 0.002 BrAC (218 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.	Concentration	Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2015.12.21 14:02:23 -06:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

DAN R LUCAS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT, INTOX 8000, INTOX EC/IR II, ASIV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/15/2015

NUMBER 250138

EXPIRES 6/15/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)