RECEIVED
By Carol Day at 4:26 pm, Feb 19, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

INTOX EC/IR II MAINTENANCE REPORT

REPORT #3

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX EC/IR II SN	NAME OF AGENCY	DATE OF INSPECTION
12828	Warrensburg P.D.	02/09/2016
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
102 B South Holden Stree Warrensburg		11:38 CST

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

[X] **DIAGNOSTIC RECORD**

- [X] BLANK CHECK
- [X] FC 1 TEMP
- [X] SRC TEMP
- [X] DET TEMP
- [X] BT TEMP
- [X] STD 2 TEMP
- [X] ETH CHECK
- [X] CO2 CHECK
- [X] FLOW CHECK
- [X] FCB CHECK
- [X] CRC COMP CHECK
- [X] CRC CAL CHECK
- [X] PRINT TEST

BREATH ANALYZER ACCURACY STANDARDS

- [ ] SIMULATOR SOLUTION
- [X] COMPRESSED ETHANOL-GAS MIXTURE
- [X] STANDARD SUPPLIER intoximeters LOT# ag421804 EXP. DATE 08/06/2016
- [ ] SIMULATOR TEMP (34°C ±0.2°C)

SIMULATOR S/N	SIMULATOR EXP DATE

[X] CALIBRATION CHECK - **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [ ] 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [X] 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- [ ] 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
0.080 g/210L	0.080 g/210L	0.080 g/210L

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS	0-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	0	0	2	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE		PRINT FULL NAME
		REYNOLDS, JEFFREY
TYPE II PERMIT NUMBER	EXPIRATION DATE	TELEPHONE NUMBER
240215	04/30/2016	(660) 747-9133

RETURN COMPLETED REPORT TO THE:

Breath Alcohol Program, Missouri Department of Health and Senior Services,
Southeast District Office, 2875 James Blvd, Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

JEFFREY E REYNOLDS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/30/2014

NUMBER 240215

EXPIRES 4/30/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (6-10)

LAB 4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator REYNOLDS, JEFFREY
Permit No 240215
Date Issued 4/30/2014 **Date Expires** 4/30/2016