



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
**INTOX EC/IR II MAINTENANCE REPORT**

REPORT #3

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX EC/IR II SN 12829	NAME OF AGENCY RAYTOWN POLICE DEPT	DATE OF INSPECTION 05/16/2016
----------------------------	---------------------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 10000 E 59TH STREET RAYTOWN	TIME OF INSPECTION 13:08 CDT
---	---------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

- DIAGNOSTIC RECORD**
- |   |  |
|---|--|
| <input checked="" type="checkbox"/> BLANK CHECK | <input checked="" type="checkbox"/> CO2 CHECK      |
| <input checked="" type="checkbox"/> FC 1 TEMP   | <input checked="" type="checkbox"/> FLOW CHECK     |
| <input checked="" type="checkbox"/> SRC TEMP    | <input checked="" type="checkbox"/> FCB CHECK      |
| <input checked="" type="checkbox"/> DET TEMP    | <input checked="" type="checkbox"/> CRC COMP CHECK |
| <input checked="" type="checkbox"/> BT TEMP     | <input checked="" type="checkbox"/> CRC CAL CHECK  |
| <input checked="" type="checkbox"/> STD 2 TEMP  | <input checked="" type="checkbox"/> PRINT TEST     |
| <input checked="" type="checkbox"/> ETH CHECK   |  |

**BREATH ANALYZER ACCURACY STANDARDS**

SIMULATOR SOLUTION       COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER      INTOXIMETERS      LOT# AG607103      EXP. DATE 03/11/2018

<input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C)	SIMULATOR S/N	SIMULATOR EXP DATE
---	---------------	--------------------

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 @ 0.101 g/210L	TEST 2 @ 0.101 g/210L	TEST 3 @ 0.101 g/210L
-----------------------	-----------------------	-----------------------

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:**

REFUSALS	1	0-.04	5	.05-.09	0	.10-.14	2	.15-.19	0	OVER .19	2
----------	---	-------	---	---------	---	---------	---	---------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Monthly Maintenance

**INSPECTING OFFICER**

SIGNATURE <i>Kevin Sheets</i>	PRINT FULL NAME kevin sheets
TYPE II PERMIT NUMBER 260194	TELEPHONE NUMBER ( 816 ) 737-6020
EXPIRATION DATE 04/26/2018	

**RETURN COMPLETED REPORT TO THE:**

Breath Alcohol Program, Missouri Department of Health and Senior Services,  
Southeast District Office, 2875 James Blvd, Poplar Bluff, MO 63901



Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-3100  
Fax: (314) 533-7328

## Certificate of Analysis

Customer Name  
*Exclusive Supplier*  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

Test Date: 14-Mar-2016

Lot # AG607103 Model 108cadd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
11-Mar-2018	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Analyst: \_\_\_\_\_

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01