



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM

**INTOX EC/IR II MAINTENANCE REPORT**

REPORT #3

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX EC/IR II SN 12859	NAME OF AGENCY Lincoln County SO	DATE OF INSPECTION 04/08/2016
----------------------------	-------------------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 65 Business Park Drive Troy	TIME OF INSPECTION 15:45 CDT
---	---------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> <b>DIAGNOSTIC RECORD</b>	
<input checked="" type="checkbox"/> BLANK CHECK	<input checked="" type="checkbox"/> CO2 CHECK
<input checked="" type="checkbox"/> FC 1 TEMP	<input checked="" type="checkbox"/> FLOW CHECK
<input checked="" type="checkbox"/> SRC TEMP	<input checked="" type="checkbox"/> FCB CHECK
<input checked="" type="checkbox"/> DET TEMP	<input checked="" type="checkbox"/> CRC COMP CHECK
<input checked="" type="checkbox"/> BT TEMP	<input checked="" type="checkbox"/> CRC CAL CHECK
<input checked="" type="checkbox"/> STD 2 TEMP	<input checked="" type="checkbox"/> PRINT TEST
<input checked="" type="checkbox"/> ETH CHECK	

**BREATH ANALYZER ACCURACY STANDARDS**

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER      intoximeters	LOT#    ag434201      EXP. DATE    12/08/2016
<input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C)	SIMULATOR S/N      SIMULATOR EXP DATE

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within +5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input checked="" type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1    0.081 g/210L	TEST 2    0.081 g/210L	TEST 3    0.081 g/210L
------------------------	------------------------	------------------------

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:**

REFUSALS    0	0-.04    3	.05-.09    1	.10-.14    1	.15-.19    0	OVER .19    1
---------------	------------	--------------	--------------	--------------	---------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

**INSPECTING OFFICER**

SIGNATURE 	PRINT FULL NAME PIRTLE, MICHAEL
TYPE II PERMIT NUMBER 250030	EXPIRATION DATE 01/20/2017
	TELEPHONE NUMBER ( 636 ) 528-8546

**RETURN COMPLETED REPORT TO THE:**

Breath Alcohol Program, Missouri Department of Health and Senior Services,  
Southeast District Office, 2875 James Blvd, Poplar Bluff, MO 63901



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
**BLOOD ALCOHOL TEST REPORT - INTOX EC/IR II**

FORM #13

LOCATION OF INSTRUMENT Lincoln County SO 65 Business Park Drive Troy		INSTRUMENT SERIAL NUMBER 12859	DATE OF TEST 04/08/2016	TIME OF TEST 15:47 CDT
SUBJECT NAME N/A, N/A			DATE OF BIRTH 01/01/1981	
SEX M	SUBJECT DRIVER'S LICENSE NUMBER 1234		STATE MO	
ARRESTING OFFICER PIRTLE, MICHAEL		ARRESTING OFFICER ID 490		
OPERATOR PIRTLE, MICHAEL		OPERATOR PERMIT 250030	PERMIT EXP DATE 01/20/2017	

**OPERATIONAL CHECKLIST: INTOX EC/IR II**

- 1. Examination of mouth conducted. If any substance is observed or indicated to be present, the substance observed or indicated must be removed prior to starting the 15 minute observation period.
- 2. Subject observed for at least 15 minutes by: PIRTLE, MICHAEL No smoking, oral intake or vomiting during this time; if vomiting occurs, start over with the 15 minute observation period.
- 3. Assure that the power switch is ON and the screen is displaying "PRESS ENTER TO START".
- 4. Press the Enter button.
- 5. Enter subject and officer information.
- 6. When display reads "Please Blow /R", and gives audible beep, insert mouthpiece and take the subject's breath sample.

**SUBJECT TEST RESULTS**

Test	g/210L	Time	Smpl #	Durn (sec)	Vol (cc)	Time
DIAG	Pass	15:48				
BLK	0.000	15:48	1	3.17	1613	15:49
SUBJ	0.000	15:49				
BLK	0.000	15:50				

COMMENTS

BLANK TEST

**CERTIFICATION BY OPERATOR**

BAC

**0.000 g/210L**

As set forth in the rules promulgated by the Department of Health and Senior Services related to the determination of blood alcohol by breath analysis, I certify that:

- 1. There was no deviation from the procedure approved by the department.
- 2. To the best of my knowledge the instrument was functioning properly.
- 3. I am authorized to operate the instrument.
- 4. No radio transmission occurred inside the room where and when this test was being conducted.

SIGNATURE OF OPERATOR

DATE

*Got. [Signature]* 490

4-8-16

WITNESS (IF ANY)

DATE