



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

REPORT #0

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN #201217	NAME OF AGENCY Crystal City Police Department	DATE OF INSPECTION 02/12/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 130 Mississippi Avenue, Crystal City		TIME OF INSPECTION 4:00 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 02/12/2016 16:24
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 51 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc.	LOT # 15120 EXP. DATE 04/29/2017
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C	SIMULATOR SN DR3772 EXP. DATE 05/18/2016

CALIBRATION CHECK -- (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → .100	TEST 2 → .099	TEST 3 → .099
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(.0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	OVER .19 0
------------	------------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

<b>INSPECTING OFFICER</b>	
SIGNATURE Sgt. Jeffery S. Wynn #204	PRINT FULL NAME Sgt. Jeffery S. Wynn, DSN-204
TYPE II PERMIT NUMBER/EXPIRATION DATE 240141 04/03/2016	TELEPHONE NUMBER (636) 937-4601

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 4, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is April 29, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Certified Reference Standard lot number EN08051301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

Face This Side Down - This Edge In First

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
CRYSTAL CITY POLICE DEPARTMENT

STATE OF MISSOURI  
CRYSTAL CITY POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 801217  
08/18/16  
14184

BAC DATA MASTER SERIAL NUMBER 801217  
08/18/16

--- BACMASTER CHECK ---

TESTING OFFICER:

MYNRY, JEFFERY/S

OFFICER I.D.# 204

PERMIT NUMBER: 240141

EXPIRATION DATE: 04/03/16

MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

COMPUTER: OKAY

PROGRAM 104-00 0000: OKAY

DEVIANCE

SAMPLE NUMBER: 516

FLOW DETECTOR: OKAY

FUMP

HIGH SPEED: OKAY

DETECTURE: OKAY

CH TEST: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

BLANK TEST	.000	16127
INTERNAL STANDARD	VERIFIED	16127
EXTERNAL STANDARD	.100	16128
BLANK TEST	.000	16128
EXTERNAL STANDARD	.699	16129
BLANK TEST	.000	16130
EXTERNAL STANDARD	.099	16130
BLANK TEST	.000	16131

N = 3

SD = .1

AVG. = .0000

PRINTER TEST  
11/22/11 12:44 - 1/01/12 20567891 | C=1700000000  
BY JOLINCE/BESTUMAKY21 | C=1700000000 | mode | g | h | i | j | k | l | m | n | o  
port # 1000000000 | 11/22/11

Operator Signature

*[Signature]* #204 Type II

2203-02

Operator Signature

*[Signature]* #204 Type II

22

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
CRYSTAL CITY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201217  
02/02/16

ARREST TIME: 16:00  
SUBJECT NAME:  
TEST #01  
DOB: 01/02/04 SEX: F  
STATE-D.L.# MO-A123456789  
ARRESTING OFFICER:  
WYNN/JEFFERY/S  
OFFICER I.D.: 004  
TESTING OFFICER:  
WYNN/JEFFERY/S  
OFFICER I.D.: 004  
PERMIT NUMBER: 040141  
EXPIRATION DATE: 04/03/16  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BAC TEST	.006	16:04
INTERNAL STANDARD	VERIFIED	16:04
RADIO INTERFERENCE		

Operator Signature *Ed J. H. 2004 Type II*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**

**JEFFERY S WYNN**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/3/2014

NUMBER 240141

EXPIRES 4/3/2016

MO 680-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (06-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator WYNN, JEFFERY  
Permit No 240141  
Date Issued 4/3/2014 Date Expires 4/3/2016