



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN #201217	NAME OF AGENCY Crystal City Police Department	DATE OF INSPECTION 04/25/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 130 Mississippi Avenue, Crystal City		TIME OF INSPECTION 1:00 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 04/25/2016 13:03
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc.	LOT # 15120 EXP. DATE 04/29/2017
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C	SIMULATOR SN DR3772 EXP. DATE 05/18/2016

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → .097	TEST 2 → .096	TEST 3 → .096
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 1	OVER .19 0
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Sgt. Jeffery S. Wynn, DSN-204
TYPE II PERMIT NUMBER/EXPIRATION DATE 260175 04/04/2018	TELEPHONE NUMBER (636) 937-4601

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 4, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is April 29, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MICHIGAN
CRYSTAL CITY POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER PRINTED
04/28/13
13193

--- DIAGNOSTIC CHECK ---

- COMPILERS COPY
- PROGRAM COPY BY 06/28/13 COPY
- HEATERS
- SAMPLE CHAMBERS 250
- SLOW VELOCITY COPY
- 2000
- HIGH SPEED COPY
- DETECTORS COPY
- FILTERS COPY
- QUARTZ STANDARD COPY
- LOW INJECTION COPY

PRINTED ISSY

Crystal City Police Department
1000 Michigan Street
Crystal City, Michigan 48128

Operator Signature

[Signature] #204 Type II

2209-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MICHIGAN
CRYSTAL CITY POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER PRINTED
04/28/13

TESTING OFFICER:

BYRON JEFFERY'S

OFFICER I.D.# 204

PERMIT NUMBER: 266175

EXPIRATION DATE: 04/04/18

MISCELLANEOUS DATA:

--- SUPERVISOR NOTE ---

BLANK TEST	.000	13:06
INTERNAL STANDARD	VERIFIED	13:06
EXTERNAL STANDARD	.007	13:07
BLANK TEST	.000	13:08
EXTERNAL STANDARD	.006	13:08
BLANK TEST	.000	13:09
EXTERNAL STANDARD	.006	13:09
BLANK TEST	.000	13:10

n = 3

StD. = .1

AVG. = .000

Operator Signature

[Signature] #204 Type II

221

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
CRYSTAL CITY POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201217
04/25/10

ARREST TIME: 12:40
SUBJECT NAME:
TEST/RTI
DOB: 01/27/45 SEX: M
STREET: 1111 W. BENTLEY ST
ARRESTING OFFICER:
OFFICER JEFFERY S
OFFICER L.H. # 204
TESTING OFFICER:
OFFICER JEFFERY S
OFFICER L.D. # 204
EVIDENCE NUMBER: 060175
EXPIRATION DATE: 06-30-10
MISCELLANEOUS INFO:

----- PRINTED NAME TEST -----

BLIND TEST	0.000	13:14
1P CEREBR 5 CONDENSE	0.000	13:15
RIGHT INTER-OCCIPIT		

Operator Signature

[Handwritten Signature] Type II



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

JEFFERY S WYNN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/4/2016

NUMBER 260175

EXPIRES 4/4/2018

MO 580-Q771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (06-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WYNN, JEFFERY
Permit No 260175
Date Issued 4/4/2016 Date Expires 4/4/2018