



RECEIVED
By Carol Day at 2:39 pm, Jun 02, 2016

DATAMASTER MAINTENANCE REPORT

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 202091	NAME OF AGENCY Directorate of Emergency Services	DATE OF INSPECTION 06/02/2016
-------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Bldg 1000, Fort Leonard Wood	TIME OF INSPECTION 1358
--	----------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 06/02/16 1358
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER GUTH LABS INC LOT # 14220 EXP. DATE 09/24/2016

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN DR5376 EXP. DATE 09/10/2016

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .101	TEST 2 .101	TEST 3 .101
-------------	-------------	-------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	OVER .19 0
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

TESTED AND CERTIFIED AS WITHIN GUIDELINES ESTABLISHED BY DHSS.

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME ROBERT ISHMAEL
---------------	-----------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 260085 02/22/2018	TELEPHONE NUMBER (573) 596-0585
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

FORT LEONARD WOOD
DIRECTORATE OF EMERGENCY SERVICES

BAC DATAMASTER SERIAL NUMBER 202091

06/02/16

13:58

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS

SAMPLE CHAMBER: 50c

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY


QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopq
rstuvwxyz{|}~

Operator Signature



Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

FORT LEONARD WOOD
DIRECTORATE OF EMERGENCY SERVICES

BAC DATAMASTER SERIAL NUMBER 202091
06/02/16

TESTING OFFICER:
ISHMAEL/ROBERT
OFFICER I.D.: 18740
PERMIT NUMBER: 260085
EXPIRATION DATE: 02/22/18
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	14:01
INTERNAL STANDARD	VERIFIED	14:01
EXTERNAL STANDARD	.101	14:01
BLANK TEST	.000	14:02
EXTERNAL STANDARD	.101	14:02
BLANK TEST	.000	14:03
EXTERNAL STANDARD	.101	14:03
BLANK TEST	.000	14:04

N = 3
SIM. = .1
AVG. = .101

Operator Signature



Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

FORT LEONARD WOOD
DIRECTORATE OF EMERGENCY SERVICES

BAC DATAMASTER SERIAL NUMBER 202091
06/02/16

ARREST TIME: 01:00
SUBJECT NAME:
NA
DOB: 01/01/01 SEX: M
STATE/D.L.: NA/NA
ARRESTING OFFICER:
NA
OFFICER I.D.: NA
TESTING OFFICER:
ISHMAEL/ROBERT
OFFICER I.D.: 18740
PERMIT NUMBER: 260085
EXPIRATION DATE: 02/22/18
MISCELLANEOUS DATA:
RFI/TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	14:06
INTERNAL STANDARD	VERIFIED	14:06
RADIO INTERFERENCE		

Operator Signature _____

2208-02



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ROBERT ISHMAEL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/22/2016

NUMBER 260085

EXPIRES 2/22/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator ISHMAEL, ROBERT
Permit No 260085
Date Issued 2/22/2016 Date Expires 2/22/2018



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14220** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 25, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 24, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.