





## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14220** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 25, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 24, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FX08051301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
BAC DATAMASTER SERIAL NUMBER 283885  
04/25/16

ARREST TIME: 09:58  
SUBJECT NAME:  
DUE/JOHN  
DOB: 01/01/82 SEX: M  
STATE/D.L.: MO/123456789  
ARRESTING OFFICER:  
PATRICK/DANIEL  
OFFICER I.D.: 20  
TESTING OFFICER:  
PATRICK/DANIEL  
OFFICER I.D.: 20  
PERMIT NUMBER: 260092  
EXPIRATION DATE: 02/22/18  
MISCELLANEOUS DATA:  
MONTHLY MAINTENANCE

--- BREATH ANALYSIS ---

BLANK TEST .000 09:32  
INTERNAL STANDARD VERIFIED 09:32  
RADIO INTERFERENCE

OPERATOR SIGNATURE

\* No

REORDER ALL SUPPLIES FROM N.P.A.S.  
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
BAC DATAMASTER SERIAL NUMBER 283885  
04/25/16

ARRESTING OFFICER:  
PATRICK/DANIEL  
OFFICER I.D.: 20  
PERMIT NUMBER: 260092  
EXPIRATION DATE: 02/22/18  
MISCELLANEOUS DATA:  
MONTHLY MAINTENANCE

--- SUPERVISOR MODE ---

BLANK TEST .000 09:36  
INTERNAL STANDARD VERIFIED 09:36  
EXTERNAL STANDARD .098 09:37  
BLANK TEST .000 09:37  
EXTERNAL STANDARD .099 09:38  
BLANK TEST .000 09:38  
EXTERNAL STANDARD .099 09:39  
BLANK TEST .000 09:39

N = 3  
SIM. = .1  
AVG. = .0986

OPERATOR SIGNATURE

Card Stock No.  
60021

REORDER ALL SUPPLIES FROM N.P.A.S.  
P.O. BOX 1435, MANSFIELD, OH 44901





STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**DANIEL R PATRICK**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV W PRINTER, DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/22/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 260092

EXPIRES 2/22/2018

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator PATRICK, DANIEL  
 Permit No 260092  
 Date Issued 2/22/2016 Date Expires 2/22/2018