



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED

By Ellen Strawsine at 2:06 pm, Jan 19, 2016

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 206007	NAME OF AGENCY Louisiana P.D.	DATE OF INSPECTION 01/05/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 202 S. 3rd Street Louisiana, Missouri		TIME OF INSPECTION 6:31 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/05/2016 @18:31
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth	LOT # 15120 EXP. DATE 04/29/2017
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34 °C	SIMULATOR SN SD1451 EXP. DATE 08/10/2016

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • .099	TEST 2 • .101	TEST 3 • .101
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	1	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument Operating Within Established Limits.

INSPECTING OFFICER	
SIGNATURE <i>M. McCollister</i>	PRINT FULL NAME M. McCollister
TYPE II PERMIT NUMBER/EXPIRATION DATE 240304 07/22/2016	TELEPHONE NUMBER (573) 754-4021

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **15120** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 4, 2015**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 29, 2017** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** $\pm 3\%$.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
MARC MCCOLLISTER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2014

NUMBER 240304

EXPIRES 7/22/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MD 580-0771 (8-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MCCOLLISTER, MARC
 Permit No 240304
 Date issued 7/22/2014 Date Expires 7/22/2016

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STREET OR ALCOHOL:

3RD DISTRICTER SERIAL NUMBER 222227

1/15/15

TESTING OFFICER:

*CALIBRATION# 2

OFFICER I.D.# 111

PERMIT NUMBER: 240004

EXPIRATION DATE: 07/22/15

MISCELLANEOUS DATA:

N/A

N/A

--- SUPERVISOR MODE ---

BLANK TEST	1.000	120.24
INTERNAL STANDARD	VERIFIED	121.24
EXTERNAL STANDARD	1.025	119.74
BLANK TEST	1.000	121.20
EXTERNAL STANDARD	1.121	123.30
BLANK TEST	1.000	121.25
EXTERNAL STANDARD	1.121	124.25
BLANK TEST	1.002	121.27

--- N = 3

STN. = 11

AVG. = 1.025

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STREET OR ALCOHOL:

3RD DISTRICTER SERIAL NUMBER 222227

01/15/15

15451

--- LABORATORY CHECK ---

COMPUTER: OKAY

PROGRAM (04-27-2012): OKAY

REPEERS

REFILL OPERERS: OK

FLDM DETECTOR: OKAY

FLDM: OKAY

FLOW SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY


ALBERTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

1145728/02000001/012245572911(=)PANDORA
 HIKUMORPQRSTUWXYZ1234567890ABCDEFGHIJKL
 01234567890123456789012345678901234567890

Operator Signature



2208-02

Operator Signature



2208-02

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
DPO DATAMASTER SERIAL NUMBER 200022
01/05/16

PRETEST TIME: 19:00
SUBJECT NAME:

DOB/DOB/A

DOB: 09/05/69 SEX: M

STATE/D.L.# NO/000000000000000000

PRETESTING OFFICER:

MOBILIST/NO/A

OFFICER (A.D.# 01)

TESTING OFFICER:

MOBILIST/NO/A

OFFICER (A.D.# 01)

PERMIT NUMBER: 040324

EXPIRATION DATE: 07/02/16

MISCELLANEOUS DATA:

RPI TEST

N/A

OPERATOR SIGNATURE

BLANK TEST

INTERNAL STORAGE

RADIO INTERFERENCE

DATE

VERIFIED

19:45

Operator Signature

