MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 8:06 am, Jul 20, 2016

REPORT #6

Complete this report at the time of the regular monthly preventive [illegible]
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
206007	Louisiana P.D.	06/28/2016

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
202 S. 3rd Street Louisiana, Missouri	11:54 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 06/28/216 @ 23:54

☑ COMPUTER	☑ DETECTOR
☑ PROGRAM	☑ FILTERS
☑ HEATERS SAMPLE CHAMBER 50 °C	☑ QUARTZ STANDARD
☑ FLOW DETECTOR	☑ CALIBRATION
☑ PUMP HIGH SPEED	☑ PRINTER

☑ INDICATOR LIGHTS

☑ SIMULATOR SOLUTION SUPPLIER Guth Laboratories LOT # 16040 EXP. DATE 01/20/2018

☑ SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD1451 EXP. DATE 04/27/2017

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☛ .097	TEST 2 ☛ .097	TEST 3 ☛ .098

☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	0	0	0	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument Operating Within Established Limits.

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
▸ M. McCollister	M. McCollister
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
240304 07/22/2016	(573) 754-4021

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **16040** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **January 22, 2016,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1213%** (w/vol) ethyl alcohol. The expiration date for this lot number is **January 20, 2018** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
*Testing was conducted using Cerilliant Reference Standard lot number **FN08051301** whose values are traceable to NIST.*
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

MARC MCCOLLISTER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2014

NUMBER 240304

EXPIRES 7/22/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R5-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MCCOLLISTER, MARC
Permit No 240304
Date Issued 7/22/2014 Date Expires 7/22/2016

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER [illegible]
06/28/16

TESTING OFFICER:
MCCALLISTER, MARCUS
OFFICER [illegible]
PERMIT NUMBER [illegible]
EXPIRATION [illegible]
[illegible]
[illegible]
[illegible]

[illegible]

BLANK TEST	[illegible]	[illegible]
INTERNAL STANDARD	[illegible]	[illegible]
EXTERNAL STANDARD	[illegible]	[illegible]
BLANK TEST	[illegible]	[illegible]
EXTERNAL [illegible]	[illegible]	[illegible]
BLANK TEST	[illegible]	[illegible]
EXTERNAL [illegible]	[illegible]	[illegible]
BLANK TEST	[illegible]	[illegible]

[illegible]
[illegible]
[illegible]

Operator Signature [signature]

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER [illegible]
06/28/16
[illegible]

[illegible] DIAGNOSTIC [illegible]

COMPUTER:	[illegible]
PROGRAM (04 [illegible]):	[illegible]
HEATERS	
SAMPLE [illegible]:	[illegible]
[illegible]	[illegible]
[illegible]	
[illegible]	[illegible]
[illegible]	[illegible]
[illegible]	[illegible]
[illegible]	[illegible]
[illegible]	[illegible]

[illegible]
[illegible]
[illegible]
[illegible]

Operator Signature [signature]

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

Operator Signature _____

2208-02