



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

INTOX DMT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX DMT SN 500062	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 05/14/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 511 W. CLAY, ROCK PORT, MO		TIME OF INSPECTION 15:30:08

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

DIAGNOSTIC RECORD

DATE AND TIME <u>05/14/2016 15:30:11</u>	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTER 1
<input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.8°C</u>	<input checked="" type="checkbox"/> FILTER 2
<input checked="" type="checkbox"/> BREATH TUBE <u>43.5°C</u>	<input checked="" type="checkbox"/> FILTER 3
<input checked="" type="checkbox"/> PUMP	<input checked="" type="checkbox"/> INTERNAL STANDARD

BREATH ANALYZER ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR STANDARD	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
----------------------------------------------------	---------------------------------------------------------------------------

STANDARD SUPPLIER INTOXIMETERS LOT # AG604101 EXP. DATE 02/10/2018

SIMULATOR TEMP (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: 0.097	TEST 2: 0.097	TEST 3: 0.097
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS: 0	0-.04: 42	.05-.09: 0	.10-.14: 1	.15-.19: 1	OVER .19: 0
--------------------	------------------	-------------------	-------------------	-------------------	--------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

maintenance following recalibration

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME JACK W COTTER
---------------	-----------------------------------------

TYPE II PERMIT NUMBER 260074	EXPIRATION DATE 02/18/2018	TELEPHONE NUMBER 816-379-5161
----------------------------------------	--------------------------------------	-----------------------------------------

RETURN COMPLETED REPORT TO THE **Breath Alcohol Program, MO Department of Health and Senior Services
Southeast District Office
2875 James Blvd, Poplar Bluff, MO 63901**

CALIBRATION FACTORS

Missouri State Highway Patrol
INTOX dmt: 500062

Date: 05/14/2016
Time: 14:51:22

OPERATOR NAME:
JACK W COTTER
PERMIT NUMBER: 260074
EXPIRATION DATE: 02/18/2018

LOT #: 15001
SUPPLIER: REPCO
EXPIRATION: 05/20/2017

Ca = 0.1000
ADJ = 0.973067 0.800 <= ADJ < 1.200
b1 = 0.0010 0.0000 <= b1 < 0.0040
b2 = 0.0056 0.0010 <= b2 < 0.0100
b3 = 0.0003 0.0000 <= b3 < 0.0040
Xq = 0.0811 0.0500 <= Xq < 0.2500
a21 = 1.084320 1.050 <= a21 < 1.300
a31 = 0.405110 0.300 <= a31 < 0.800

