



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
INTOX DMT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

| | | |
|--|---|----------------------------------|
| INTOX DMT SN 500261 | NAME OF AGENCY Lake Ozark Police Dept. | DATE OF INSPECTION 05/15/2016 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 3162 Bagnell Dam Boulevard, Lake Ozark | | TIME OF INSPECTION 20:51:14 |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

| | |
|--|---|
| <input checked="" type="checkbox"/> DIAGNOSTIC RECORD | |
| DATE AND TIME <u>05/15/2016 20:51:16</u> | <input checked="" type="checkbox"/> DETECTOR |
| <input checked="" type="checkbox"/> PROGRAM | <input checked="" type="checkbox"/> FILTER 1 |
| <input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.9°C</u> | <input checked="" type="checkbox"/> FILTER 2 |
| <input checked="" type="checkbox"/> BREATH TUBE <u>48.1°C</u> | <input checked="" type="checkbox"/> FILTER 3 |
| <input checked="" type="checkbox"/> PUMP | <input checked="" type="checkbox"/> INTERNAL STANDARD |

BREATH ANALYZER ACCURACY STANDARDS

| | |
|---|--|
| <input type="checkbox"/> SIMULATOR STANDARD | <input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |
|---|--|

| | | |
|---|-----------------------|-----------------------------|
| <input checked="" type="checkbox"/> STANDARD SUPPLIER <u>INTOXIMETERS</u> | LOT # <u>AG600504</u> | EXP. DATE <u>01/05/2018</u> |
| <input type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) _____ | SIMULATOR SN _____ | SIMULATOR EXP DATE _____ |

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

| |
|--|
| <input type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE |
| <input checked="" type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE |
| <input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE |

| | | |
|---------------|---------------|---------------|
| TEST 1: 0.077 | TEST 2: 0.077 | TEST 3: 0.077 |
|---------------|---------------|---------------|

PERFORM R.F.I. TEST

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

| | | | | | |
|-------------|----------|----------|----------|----------|------------|
| REFUSALS: 0 | 0-04: 39 | 05-09: 3 | 10-14: 1 | 15-19: 2 | OVER 19: 2 |
|-------------|----------|----------|----------|----------|------------|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

INSPECTING OFFICER

| | | |
|---------------------------------|------------------------------------|----------------------------------|
| SIGNATURE | PRINT FULL NAME BRIAN N JARRETT | |
| TYPE II PERMIT NUMBER 240393 | EXPIRATION DATE 11/13/2016 | TELEPHONE NUMBER 573-365-5371 |

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, MO Department of Health and Senior Services
Southeast District Office
2875 James Blvd, Poplar Bluff, MO 63901



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 7-Jan-2016

Lot # AG600504 Model 108cacd

| | | | |
|-------------------------|-------------------------|-------------------------|---|
| <u>Exp. Date</u> | <u>Cyl. Type</u> | <u>Component</u> | <u>Certified Concentration</u> |
| 5-Jan-2018 | 108 | Ethanol Nitrogen | 0.080 ± 0.002 BrAC (208 ppm) Balance |

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

| <u>Serial No.</u> | <u>Concentration</u> | <u>Serial No.</u> | <u>Concentration</u> |
|--------------------------|-----------------------------|--------------------------|-----------------------------|
| EB0010581 | 391.8 ppm | EB0010603 | 392.5 ppm |
| EB0010570 | 259.8 ppm | EB0010559 | 258.9 ppm |
| EB0010285 | 209.0 ppm | EB0010595 | 208.9 ppm |
| EB0010561 | 103.7 ppm | EB0010562 | 104.9 ppm |
| EB0010681 | 52.22 ppm | EB0010579 | 52.94 ppm |

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2016.01.07 16:36:00 -06:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

BRIAN N JARRETT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/13/2014

NUMBER 240393

EXPIRES 11/13/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator JARRETT, BRIAN
Permit No 240393
Date Issued 11/13/2014 Date Expires 11/13/2016