



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**INTOX DMT MAINTENANCE REPORT**

**RECEIVED**

*By Carol Day at 9:08 am, Jul 11, 2016*

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

|                                                                                   |                                                   |                                         |
|-----------------------------------------------------------------------------------|---------------------------------------------------|-----------------------------------------|
| INTOX DMT SN<br><b>500262</b>                                                     | NAME OF AGENCY<br><b>Malden Police Department</b> | DATE OF INSPECTION<br><b>07/08/2016</b> |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br><b>112 E. Laclede Malden MO 63863</b> |                                                   | TIME OF INSPECTION<br><b>15:15:21</b>   |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

**DIAGNOSTIC RECORD**

|                                                                  |                                                       |
|------------------------------------------------------------------|-------------------------------------------------------|
| DATE AND TIME <u>07/08/2016 15:15:24</u>                         | <input checked="" type="checkbox"/> DETECTOR          |
| <input checked="" type="checkbox"/> PROGRAM                      | <input checked="" type="checkbox"/> FILTER 1          |
| <input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.8°C</u> | <input checked="" type="checkbox"/> FILTER 2          |
| <input checked="" type="checkbox"/> BREATH TUBE <u>44.6°C</u>    | <input checked="" type="checkbox"/> FILTER 3          |
| <input checked="" type="checkbox"/> PUMP                         | <input checked="" type="checkbox"/> INTERNAL STANDARD |

**BREATH ANALYZER ACCURACY STANDARDS**

SIMULATOR STANDARD       COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETER      LOT # AG523101      EXP. DATE 08/19/2017

SIMULATOR TEMP (34°C ± 0.2°C) \_\_\_\_\_      SIMULATOR SN \_\_\_\_\_      SIMULATOR EXP DATE \_\_\_\_\_

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**  
 Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: **0.097**      TEST 2: **0.096**      TEST 3: **0.096**

**PERFORM R.F.I. TEST**

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

|             |          |            |            |            |             |
|-------------|----------|------------|------------|------------|-------------|
| REFUSALS: 0 | 0-.04: 2 | .05-.09: 1 | .10-.14: 1 | .15-.19: 0 | OVER .19: 1 |
|-------------|----------|------------|------------|------------|-------------|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

Instrument was sent to Intoximeters for values dropping while performing maintenance test. Intoximeters cleaned optics/filters, cleaned fittings, replaced inner/outer orings, calibrated to factory specs

**INSPECTING OFFICER**

|                                        |                                            |
|----------------------------------------|--------------------------------------------|
| SIGNATURE<br>                          | PRINT FULL NAME<br><b>RUSSELL L MILLER</b> |
| TYPE II PERMIT NUMBER<br><b>250117</b> | EXPIRATION DATE<br><b>05/14/2017</b>       |
| TELEPHONE NUMBER                       |                                            |

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, MO Department of Health and Senior Services  
 Southeast District Office  
 2875 James Blvd, Poplar Bluff, MO 63901



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II**

**RUSSELL L MILLER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/14/2015

NUMBER 250117

EXPIRES 5/14/2017

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator MILLER, RUSSELL  
Permit No 250117  
Date Issued 5/14/2015 Date Expires 5/14/2017



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

*Exclusive Supplier*  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 19-Aug-2015

**Lot # AG523101 Model 108cadd**

|                  |                  |                     |                                      |
|------------------|------------------|---------------------|--------------------------------------|
| <b>Exp. Date</b> | <b>Cyl. Type</b> | <b>Component</b>    | <b>Certified Concentration</b>       |
| 19-Aug-2017      | 108              | Ethanol<br>Nitrogen | 0.100 ± 2% BrAC (260 ppm)<br>Balance |

**Certification Traceable to N.I.S.T. RGM Ethanol Standards:**

| <u>Serial No.</u> | <u>Concentration</u> | <u>Serial No.</u> | <u>Concentration</u> |
|-------------------|----------------------|-------------------|----------------------|
| EB0010581         | 391.8 ppm            | EB0010603         | 392.5 ppm            |
| EB0010570         | 259.8 ppm            | EB0010559         | 258.9 ppm            |
| EB0010285         | 209.0 ppm            | EB0010595         | 208.9 ppm            |
| EB0010561         | 103.7 ppm            | EB0010562         | 104.9 ppm            |
| EB0010681         | 52.22 ppm            | EB0010579         | 52.94 ppm            |

**Analytical Method:** NDIR

Digitally signed by Quality Control  
Date: 2015.08.24 15:05:58 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

**Analyst:**

Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 2989.01**