RECEIVED
By Carol Day at 8:09 am, Feb 09, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

INTOX DMT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX DMT SN	NAME OF AGENCY	DATE OF INSPECTION
500292	Cabool Police Department	02/08/2016

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
510 Cedar Street, Cabool, MO.	06:20:25

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

☒ DIAGNOSTIC RECORD

DATE AND TIME 02/08/2016 06:20:27

☒ DETECTOR
☒ PROGRAM
☒ FILTER 1
☒ SAMPLE CHAMBER 48.8°C
☒ FILTER 2
☒ BREATH TUBE 47.6°C
☒ FILTER 3
☒ PUMP
☒ INTERNAL STANDARD

BREATH ANALYZER ACCURACY STANDARDS

☐ SIMULATOR STANDARD
☒ COMPRESSED ETHANOL-GAS MIXTURE

☒ STANDARD SUPPLIER INTOXIMETERS LOT # AG423202 EXP. DATE 08/20/2016

☐ SIMULATOR TEMP (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

☒ CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

☒ 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
☐ 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
☐ 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: 0.098	TEST 2: 0.098	TEST 3: 0.098

☒ PERFORM R.F.I. TEST

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS: 1	0-.04: 0	.05-.09: 0	.10-.14: 1	.15-.19: 0	OVER .19: 0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
	WALTER L DARTER

TYPE II PERMIT NUMBER	EXPIRATION DATE	TELEPHONE NUMBER
250108	05/14/2017	417-962-3993

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, MO Department of Health and Senior Services
Southeast District Office
2875 James Blvd, Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 20-Aug-2014

Lot # AG423202

Exp. Date	Cyl. Type	Component	Certified Concentration
20-Aug-2016	108	Ethanol	0.100 ± 2% BrAC (260 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.	Concentration	Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2014.08.22 11:07:29 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

WALTER L DARTER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/14/2015

NUMBER 250108

EXPIRES 5/14/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R5-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator DARTER, WALTER
Permit No 250108
Date Issued 5/14/2015 **Date Expires** 5/14/2017