



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
**INTOX DMT MAINTENANCE REPORT**

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX DMT SN 500292	NAME OF AGENCY Cabool Police Department	DATE OF INSPECTION 04/04/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 510 Cedar Street, Cabool, MO.		TIME OF INSPECTION 06:01:47

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

**DIAGNOSTIC RECORD**

DATE AND TIME <u>04/04/2016 06:01:49</u>	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTER 1
<input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.7°C</u>	<input checked="" type="checkbox"/> FILTER 2
<input checked="" type="checkbox"/> BREATH TUBE <u>46.7°C</u>	<input checked="" type="checkbox"/> FILTER 3
<input checked="" type="checkbox"/> PUMP	<input checked="" type="checkbox"/> INTERNAL STANDARD

**BREATH ANALYZER ACCURACY STANDARDS**

SIMULATOR STANDARD       COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS      LOT # AG423202      EXP. DATE 08/20/2016

SIMULATOR TEMP (34°C ± 0.2°C) \_\_\_\_\_      SIMULATOR SN \_\_\_\_\_      SIMULATOR EXP DATE \_\_\_\_\_

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**  
Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: 0.098      TEST 2: 0.098      TEST 3: 0.098

**PERFORM R.F.I. TEST**

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS: 1      0-.04: 0      .05-.09: 0      .10-.14: 1      .15-.19: 1      OVER .19: 0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

Tested within DHSS specs

**INSPECTING OFFICER**

SIGNATURE      PRINT FULL NAME  
WALTER L DARTER

TYPE II PERMIT NUMBER 250108      EXPIRATION DATE 05/14/2017      TELEPHONE NUMBER 417-962-3993

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, MO Department of Health and Senior Services  
Southeast District Office  
2875 James Blvd, Poplar Bluff, MO 63901



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 20-Aug-2014

**Lot #** AG423202

**Exp. Date**

20-Aug-2016

**Cyl. Type**

108

**Component**

Ethanol  
Nitrogen

**Certified Concentration**

0.100 ± 2% BrAC (260 ppm)  
Balance

**Certification Traceable to N.I.S.T. RGM Ethanol Standards:**

**Serial No.**

EB0010581  
EB0010570  
EB0010285  
EB0010561  
EB0010681

**Concentration**

391.8 ppm  
259.8 ppm  
209.0 ppm  
103.7 ppm  
52.22 ppm

**Serial No.**

EB0010603  
EB0010559  
EB0010595  
EB0010562  
EB0010579

**Concentration**

392.5 ppm  
258.9 ppm  
208.9 ppm  
104.9 ppm  
52.94 ppm

**Analytical Method:**

NDIR

Digitally signed by Quality Control  
Date: 2014.08.22 11:07:29 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

**Analyst:**

Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 2989.01**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**WALTER L DARTER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/22/2014

NUMBER 240158

EXPIRES 4/22/2016

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator **DARTER, WALTER**  
Permit No **240158**  
Date Issued **4/22/2014** Date Expires **4/22/2016**