MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED
By Carol Day at 3:03 pm, Apr 12, 2016

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance inspection or when instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER	LOCATION OF INSTRUMENT	DATE OF INSPECTION	TIME OF INSPECTION
80-005829	JASPER CO. SHERIFF	04/08/2016	02:09

CALIBRATION CHECK RESULTS

Test	g/210L	Time
Air Blank	0.000	02:11
Cal Check	0.080	02:11
Air Blank	0.000	02:11
Cal Check	0.080	02:12
Air Blank	0.000	02:12
Cal Check	0.080	02:13
Air Blank	0.000	02:13

Pass

CALIBRATION CHECK SUMMARY

STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
DRY	AG502603	01/26/2017
SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
N/A	N/A	N/A
STANDARD VALUE	STANDARD SUPPLIER	
0.080	INTOXIMETERS	

CALIBRATION CHECK RESULT 1: 0.080

CALIBRATION CHECK RESULT 2: 0.080

CALIBRATION CHECK RESULT 3: 0.080

MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)
0.0%	0.000

DIAGNOSTIC TEST RESULTS

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temperature Regulation Test	Pass

Pass

RFI TEST RESULTS

Test	g/210L	Time
Air Blank	0.000	02:14
Subject Test	RFI*	02:14
Air Blank	0.000	02:14

*RFI Detect

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
1	0	2	0	1	1

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

IS WORKING WITHIN MODHSS STANDARDS

INSPECTING OFFICER

SIGNATURE	PRINT NAME
Cpl. Calvin #349	CHRISTOPHER CALVIN

TYPE II PERMIT NUMBER	EXPIRATION DATE	TELEPHONE NUMBER
250286	02/25/2018	4173588177

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

CHRISTOPHER CALVIN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/25/2016

NUMBER 250286

EXPIRES 2/25/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator CALVIN, CHRISTOPHER
Permit No 250286
Date Issued 2/25/2016 **Date Expires** 2/25/2018

Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 27-Jan-2015

Lot # AG502603

Exp. Date	Cyl. Type	Component	Certified Concentration
26-Jan-2017	108	Ethanol	0.080 ± 0.002 BrAC (208 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.	Concentration	Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2015.01.27 10:55:17 -06:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01