

**RECEIVED**

By Carol Day at 12:31 pm, Aug 09, 2016



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005841		LOCATION OF INSTRUMENT KCMO POLICE DEPT.		DATE OF INSPECTION 08/02/2016		TIME OF INSPECTION 14:51	
<b>CALIBRATION CHECK RESULTS</b>				<b>CALIBRATION CHECK SUMMARY</b>			
Test	g/210L	Time		STANDARD TYPE DRY	STANDARD LOT # 01316080A2	STANDARD EXPIRATION DATE 02/05/2018	
Air Blank	0.000	14:53		SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A	
Cal Check	0.078	14:53		STANDARD VALUE 0.080	STANDARD SUPPLIER CMI		
Air Blank	0.000	14:54		CALIBRATION CHECK RESULT 1 0.078			
Cal Check	0.078	14:54		CALIBRATION CHECK RESULT 2 0.078			
Air Blank	0.000	14:55		CALIBRATION CHECK RESULT 3 0.078			
Cal Check	0.078	14:55		MAXIMUM DEVIATION (MUST BE WITHIN 5%) 2.5%		SPREAD (MUST BE .005 OR LESS) 0.000	
Air Blank	0.000	14:56					
<b>Pass</b>							
<b>DIAGNOSTIC TEST RESULTS</b>				<b>RFI TEST RESULTS</b>			
Voltage/Current Test		Pass		Test	g/210L	Time	
RAM Test		Pass		-----			
EEPROM Checksum Test		Pass		Air Blank	0.000	14:56	
Real Time Clock Test		Pass		Subject Test	RFI*	14:56	
DSP Test		Pass		Air Blank	0.000	14:57	
Analytical Stability Test		Pass		*RFI Detect			
Modem Test		Pass					
Temperature Regulation Test		Pass					
<b>Pass</b>				<b>Pass</b>			
<b>NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT</b>							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	0	0	0	0	0	0	
List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).							
<b>INSPECTING OFFICER</b>							
SIGNATURE 				PRINT NAME OFFICER TIM FILLPOT #4162			
TYPE II PERMIT NUMBER 250125			EXPIRATION DATE 06/18/2017			TELEPHONE NUMBER 8164828195	



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**TIMOTHY L FILLPOT**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/8/2015  
NUMBER 250125  
EXPIRES 6/8/2017

\_\_\_\_\_  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY  
  
\_\_\_\_\_  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (16-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator FILLPOT, TIMOTHY  
Permit No 250125  
Date Issued 6/8/2015 Date Expires 6/8/2017



ISO/IEC 17025:2005 Accredited Laboratory

**Certificate of Analysis**

**Certificate ID:** 8789  
**Part #:** BAC105L080T  
**Cylinder Size:** 105L  
**Lot Number:** 01316080A2  
**Expiration:** 2/5/2018

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

**Contents:** 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	288 ppm	+/- 0.002 or 2% BAC whichever is greater	NDIR
Nitrogen	Balance		

\*NIST Standard Reference Material  
Cylinder No. CC14290 / Job No. 09160202  
Certified 212.8 µmol/mol Ethanol in Nitrogen  
for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

02/04/16  
Date

**Distributed by:**

**CMI Inc.**  
316 East Ninth Street  
Owensboro, KY 42303  
Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)

