

**RECEIVED**

By Carol Day at 12:30 pm, Apr 13, 2016



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005845		LOCATION OF INSTRUMENT KANSAS CITY PD		DATE OF INSPECTION 03/31/2016		TIME OF INSPECTION 14:03	
<b>CALIBRATION CHECK RESULTS</b>				<b>CALIBRATION CHECK SUMMARY</b>			
Test	g/210L	Time		STANDARD TYPE DRY	STANDARD LOT # 05514080A1	STANDARD EXPIRATION DATE 04/01/2016	
Air Blank	0.000	14:05		SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A	
Cal Check	0.080	14:05		STANDARD VALUE 0.080	STANDARD SUPPLIER CMI		
Air Blank	0.000	14:06		CALIBRATION CHECK RESULT 1 0.080			
Cal Check	0.080	14:06		CALIBRATION CHECK RESULT 2 0.080			
Air Blank	0.000	14:07		CALIBRATION CHECK RESULT 3 0.080			
Cal Check	0.080	14:07		MAXIMUM DEVIATION (MUST BE WITHIN 5%) 0.0%		SPREAD (MUST BE .005 OR LESS) 0.000	
Air Blank	0.000	14:07		<b>Pass</b>			

<b>DIAGNOSTIC TEST RESULTS</b>				<b>RFI TEST RESULTS</b>		
Voltage/Current Test	Pass	Test	g/210L	Time		
RAM Test	Pass	-----				
EEPROM Checksum Test	Pass	Air Blank	0.000	14:08		
Real Time Clock Test	Pass	Subject Test	RFI*	14:08		
DSP Test	Pass	Air Blank	0.000	14:09		
Analytical Stability Test	Pass	*RFI Detect				
Modem Test	Pass	<b>Pass</b>				
Temperature Regulation Test	Pass					

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	14	13	10	6	1	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 		PRINT NAME OFFICER TIM FILLPOT	
TYPE II PERMIT NUMBER 250125		EXPIRATION DATE 06/18/2017	
		TELEPHONE NUMBER 8164828195	



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2183 • Fax: 217-243-7634 • www.ilmo-products.com

### Certificate of Analysis

**Certificate ID:** 6358  
**Part #:** BAC105L080T  
**Cylinder Size:** 105L  
**Lot Number:** 05514080A1  
**Expiration:** 4/1/2016

**0.080 BAC** (For the calibration of instruments used to determine breath alcohol concentration)

**Contents:** 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	208 ppm	+/- 0.002 or 2%	NDIR
Nitrogen	Bal.ance	BAC whichever is greater	

\*NIST Standard Reference Material  
Cylinder No. CCI4290 / Job No. 09160202  
Certified 212.8 µmol/mol Ethanol in Nitrogen  
for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition  
and direct sunlight. Do not allow storage area to  
exceed 52 °C (125 °F).

*Jared Mathey*  
Specialty Gas Lab Tech

03/24/14  
Date



Distributed by: CMI Inc.  
316 East Ninth Street  
Owensboro, KY 42303  
Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)

ISO/IEC 17025:2005 Accredited Laboratory



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



## PERMIT TYPE II TIMOTHY L FILLPOT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs,  
and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections  
577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/8/2015

NUMBER 250125

EXPIRES 6/8/2017

*W. V. ...*  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY  
*Gail Vosterly*  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (RS-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM  
**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol  
instrument for the determination of the alcoholic content in breath form of expired air  
in Missouri.

Operator FILLPOT, TIMOTHY  
Permit No 250125  
Date Issued 6/8/2015 Date Expires 6/8/2017

BHT CAUSE