



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005848	LOCATION OF INSTRUMENT IPD BATVAN 2	DATE OF INSPECTION 03/08/2016	TIME OF INSPECTION 09:04
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
-----	-----	-----	WET	14220	09/24/2016
Air Blank	0.000	09:09	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Cal Check	0.099	09:10	34.0	DR4900	02/09/2017
Air Blank	0.000	09:10	STANDARD VALUE	STANDARD SUPPLIER	
Cal Check	0.100	09:11	0.100	GUTH LAB INC	
Air Blank	0.000	09:12	CALIBRATION CHECK RESULT 1		
Cal Check	0.099	09:12	0.099		
Air Blank	0.000	09:13	CALIBRATION CHECK RESULT 2		
Cal Check	0.099	09:13	0.100		
Air Blank	0.000	09:13	CALIBRATION CHECK RESULT 3		
			0.099		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)		SPREAD (MUST BE .005 OR LESS)
			1.0%		0.001

Pass

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		-----	-----	-----
RAM Test	Pass		Air Blank	0.000	09:13
EEPROM Checksum Test	Pass		Subject Test	RFI*	09:14
Real Time Clock Test	Pass		Air Blank	0.000	09:14
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

Pass

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	4	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER			
SIGNATURE <i>Todd Hargis</i>		PRINT NAME TODD HARGIS	
TYPE II PERMIT NUMBER 250157	EXPIRATION DATE 07/22/2017	TELEPHONE NUMBER 816 325-7293	



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14220 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 25, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is September 24, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- .3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Missouri Department of Health and Senior Services
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466
 Peter Lyskowski
 Director



Jeremiah W. (Jay) Nixon
 Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: DR4900 Manufacturer: Guth
 Model Number: 2100
 Agency: INDEPENDENCE PD
 Agency Address: 223 N MEMORIAL, INDEPENDENCE, MO 64050

NIST THERMOMETER INFORMATION

Serial Number: 093752 Bias: 0.00
 Uncertainty: 0.02
 Date of Certification: 9/8/2015 Date of Expiration: 9/8/2016

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	.00	0.02

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS


No adjustment was needed.

Date of testing: 2/9/2016
 Certification Expiration: 2/9/2017
 Simulator testing technician: D. LUCAS

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: ELLEN STRAWSINE
 Certification No: DR4900_292016

X 

DHSS BAP Scientist Approving



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
TODD W HARGIS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2015

NUMBER 250157

EXPIRES 7/22/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R16-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator HARGIS, TODD
Permit No 250157
Date Issued 7/22/2015 Date Expires 7/22/2017

IN THE STATE OF MISSOURI
COUNTY OF JACKSON

AFFIDAVIT

Before me, the undersigned authority personally appears Todd Hargis.
Who, being by me duly sworn, deposed as follows:

My name is Todd Hargis I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated.

I am the custodian of the records of *Intoxilyzer 8000* Serial number 80-005848.
Attached hereto are 4 pages of records from the **Independence Missouri Police Department**. These pages of records are kept by the **Independence Missouri Police Department** in regular course of business of the **Independence Missouri Police Department** for an employee or representative of the **Independence Missouri Police Department** with the knowledge of the act, event, condition, opinion, or diagnosis recorded to make the record or to transmit information thereof to be included in such record, and the record was made at or near the time of the act, event, condition, opinion or diagnoses. The records attached hereto are the original or exact duplicates of the original.

Todd Hargis
AFFIANT

Subscribed and sworn to me on this 14th day of March, 2016.

Samantha Morris
NOTARY PUBLIC



SAMANTHA MORRIS
My Commission Expires
January 24, 2017
Jackson County
Commission #13439145